

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AparnaParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: AMANDEEPRun Number: LB97677

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP65054
calibration std. phosphate 1 ppm	WP67390
calibration std. phosphate 0.5 ppm	WP67391
calibration std. phosphate 0.3 ppm	WP67392
calibration std. 0 ppm	WP67395
phosphate ICV std.	WP67399
phosphate CCV std.	WP67466
calibration std. phosphate 0.1 ppm	WP67393
calibration std. phosphate 0.05 ppm	WP67394
Combined reagent	WP67468
11 N sulfuric acid	WP66905
Phenolphthalein indicator	WP66907
Sodium hydroxide, 1N	WP66924

Intercept: -0.0011Slope: 0.6484Regression: 0.999864

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	08/31/2018	11:00
2	CAL2	0.05	1	50	50	0.032	08/31/2018	11:00
3	CAL3	0.10	1	50	50	0.065	08/31/2018	11:01
4	CAL4	0.30	1	50	50	0.186	08/31/2018	11:01
5	CAL5	0.50	1	50	50	0.328	08/31/2018	11:02
6	CAL6	1.00	1	50	50	0.647	08/31/2018	11:02

Analysis Method: 365.3ANALYST: AparnaParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: AMANDEEPRun Number: LB97677

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.323	0.500	08/31/2018	11:03
2	ICB		1	50	50	0.000	0.002	08/31/2018	11:03
3	CCV1	0.50	1	50	50	0.327	0.506	08/31/2018	11:04
4	CCB1		1	50	50	0.000	0.002	08/31/2018	11:04
5	PB112554BL		1	50	50	0.000	0.002	08/31/2018	11:05
6	PB112554BS	0.50	1	50	50	0.322	0.498	08/31/2018	11:05
7	J4740-01		1	50	50	0.022	0.036	08/31/2018	11:06
8	J4740-01DUP		1	50	50	0.022	0.036	08/31/2018	11:06
9	J4740-01MS	0.50	1	50	50	0.318	0.492	08/31/2018	11:07
10	J4740-01MSD	0.50	1	50	50	0.319	0.494	08/31/2018	11:07
11	J4740-02		1	50	50	0.027	0.043	08/31/2018	11:08
12	CCV2	0.50	1	50	50	0.326	0.504	08/31/2018	11:08
13	CCB2		1	50	50	0.000	0.002	08/31/2018	11:09