

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 08/31/2018

TEMP1 IN: 103 08/31/2018 14:00 TEMP OUT: 104 08/31/2018 15:15

TEMP2 IN: 104 08/31/2018 16:00 TEMP OUT: 104 09/04/2018 07:00

TEMP3 IN: 105 09/04/2018 09:00 TEMP OUT: 105 09/04/2018 10:35

TEMP4 IN: 105 09/04/2018 11:20 TEMP OUT: 105 09/04/2018 13:40

Run Number: LB97693

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB97693BL	LB97693BL	1.4074	1.4074	100	1.4074	1.4074	1.4074	0.0000	0
2	LB97693BS	LB97693BS	1.4037	1.4037	100	1.4540	1.4540	1.4540	0.0503	503
3	J4759-01	083018-SIDI-E-D-B	1.3872	1.3872	250	1.3948	1.3948	1.3948	0.0076	30.4
4	J4759-01DUP	083018-SIDI-E-D-BDUP	1.3816	1.3816	250	1.3893	1.3893	1.3893	0.0077	30.8
5	J4759-02	083018-SIDI-E-D-S	1.3850	1.3850	300	1.3952	1.3952	1.3952	0.0102	34
6	J4759-03	083018-SIDI-E-U-B	1.3935	1.3935	300	1.4078	1.4078	1.4078	0.0143	47.7
7	J4759-04	083018-SIDI-E-U-S	1.3832	1.3832	300	1.3919	1.3919	1.3919	0.0087	29
8	J4759-05	083018-DBRI-F-D-B	1.3963	1.3963	400	1.4088	1.4088	1.4088	0.0125	31.3
9	J4759-06	083018-DBRI-F-D-M	1.3972	1.3972	400	1.4045	1.4045	1.4045	0.0073	18.3
10	J4759-07	083018-DBRI-F-D-S	1.3919	1.3919	400	1.3989	1.3989	1.3989	0.0070	17.5
11	J4759-08	083018-DBRI-F-U-B	1.3980	1.3980	300	1.4089	1.4089	1.4089	0.0109	36.3
12	J4759-09	083018-DBRI-F-U-M	1.4052	1.4052	400	1.4123	1.4123	1.4123	0.0071	17.8
13	J4759-10	083018-DBRI-F-U-S	1.3975	1.3975	400	1.4031	1.4031	1.4031	0.0056	14
14	J4759-11	083018-TIPC-F-D-B	1.3798	1.3798	400	1.3912	1.3912	1.3912	0.0114	28.5
15	J4759-12	083018-TIPC-F-D-S	1.3894	1.3894	400	1.3989	1.3989	1.3989	0.0095	23.7

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 08/31/2018

TEMP1 IN: 103 08/31/2018 14:00 TEMP OUT: 104 08/31/2018 15:15

TEMP2 IN: 104 08/31/2018 16:00 TEMP OUT: 104 09/04/2018 07:00

TEMP3 IN: 105 09/04/2018 09:00 TEMP OUT: 105 09/04/2018 10:35

TEMP4 IN: 105 09/04/2018 11:20 TEMP OUT: 105 09/04/2018 13:40

Run Number: LB97693

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	J4759-13	083018-TIPC-F-U-B	1.3975	1.3975	300	1.4091	1.4091	1.4091	0.0116	38.7
17	J4759-14	083018-TIPC-F-U-S	1.4051	1.4051	300	1.4134	1.4134	1.4134	0.0083	27.7
18	J4759-15	083018-JETT-F-D-B	1.3969	1.3969	300	1.4094	1.4094	1.4094	0.0125	41.7
19	J4759-16	083018-JETT-F-D-M	1.4120	1.4120	300	1.4236	1.4236	1.4236	0.0116	38.7
20	J4759-17	083018-JETT-F-D-S	1.3801	1.3801	400	1.3885	1.3885	1.3885	0.0084	21
21	J4759-18	083018-JETT-F-U-B	1.3994	1.3994	200	1.4144	1.4144	1.4144	0.0150	75
22	J4759-19	083018-JETT-F-U-M	1.3940	1.3940	300	1.4062	1.4062	1.4062	0.0122	40.7
23	J4759-20	083018-JETT-F-U-S	1.4025	1.4025	200	1.4041	1.4041	1.4041	0.0016	8
24	J4759-20DUP	083018-JETT-F-U-SDUP	1.3935	1.3935	200	1.3951	1.3951	1.3951	0.0016	8

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS090418

WorkList ID : 116316

Date : 9/4/2018 7:28:06 AM

2297693

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/06/2018	Water	J4759-01	TSS	Cool 4 deg C	HDR103	H51	083018-SIDI-E-D-B	08/30/2018	SM2540 D
09/06/2018	Water	J4759-02	TSS	Cool 4 deg C	HDR103	H51	083018-SIDI-E-D-S	08/30/2018	SM2540 D
09/06/2018	Water	J4759-03	TSS	Cool 4 deg C	HDR103	H51	083018-SIDI-E-U-B	08/30/2018	SM2540 D
09/06/2018	Water	J4759-04	TSS	Cool 4 deg C	HDR103	H51	083018-SIDI-E-U-S	08/30/2018	SM2540 D
09/06/2018	Water	J4759-05	TSS	Cool 4 deg C	HDR103	H51	083018-DBRI-F-D-B	08/30/2018	SM2540 D
09/06/2018	Water	J4759-06	TSS	Cool 4 deg C	HDR103	H51	083018-DBRI-F-D-M	08/30/2018	SM2540 D
09/06/2018	Water	J4759-07	TSS	Cool 4 deg C	HDR103	H51	083018-DBRI-F-D-S	08/30/2018	SM2540 D
09/06/2018	Water	J4759-08	TSS	Cool 4 deg C	HDR103	H51	083018-DBRI-F-U-B	08/30/2018	SM2540 D
09/06/2018	Water	J4759-09	TSS	Cool 4 deg C	HDR103	H51	083018-DBRI-F-U-M	08/30/2018	SM2540 D
09/06/2018	Water	J4759-10	TSS	Cool 4 deg C	HDR103	H51	083018-DBRI-F-U-S	08/30/2018	SM2540 D
09/04/2018	Water	J4759-11	TSS	Cool 4 deg C	HDR103	H51	083018-TIPC-F-D-B	08/30/2018	SM2540 D
09/04/2018	Water	J4759-12	TSS	Cool 4 deg C	HDR103	H51	083018-TIPC-F-D-S	08/30/2018	SM2540 D
09/04/2018	Water	J4759-13	TSS	Cool 4 deg C	HDR103	H51	083018-TIPC-F-U-B	08/30/2018	SM2540 D
09/04/2018	Water	J4759-14	TSS	Cool 4 deg C	HDR103	H51	083018-TIPC-F-U-S	08/30/2018	SM2540 D
09/06/2018	Water	J4759-15	TSS	Cool 4 deg C	HDR103	H51	083018-JETT-F-D-B	08/30/2018	SM2540 D
09/06/2018	Water	J4759-16	TSS	Cool 4 deg C	HDR103	H51	083018-JETT-F-D-M	08/30/2018	SM2540 D
09/06/2018	Water	J4759-17	TSS	Cool 4 deg C	HDR103	H51	083018-JETT-F-D-S	08/30/2018	SM2540 D
09/06/2018	Water	J4759-18	TSS	Cool 4 deg C	HDR103	H51	083018-JETT-F-U-B	08/30/2018	SM2540 D
09/06/2018	Water	J4759-19	TSS	Cool 4 deg C	HDR103	H51	083018-JETT-F-U-M	08/30/2018	SM2540 D
09/06/2018	Water	J4759-20	TSS	Cool 4 deg C	HDR103	H51	083018-JETT-F-U-S	08/30/2018	SM2540 D

Date/Time 09/04/18, 08:00
 Received by: AS
 Relinquished by: AS

Date/Time 09/04/18, 12:30
 Received by: AS
 Relinquished by: AS