



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 09/04/2018

TEMP1 IN: 103 09/04/2018 07:00 TEMP OUT: 104 09/04/2018 08:05

TEMP2 IN: 104 09/04/2018 09:00 TEMP OUT: 104 09/04/2018 10:05

TEMP3 IN: 105 09/04/2018 11:00 TEMP OUT: 105 09/04/2018 12:40

TEMP4 IN: 105 09/04/2018 13:30 TEMP OUT: 105 09/04/2018 15:00

Run Number: LB97694

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB97694BL	LB97694BL	1.3851	1.3851	100	1.3851	1.3851	1.3851	0.0000	0
2	LB97694BS	LB97694BS	1.3901	1.3901	100	1.4411	1.4411	1.4411	0.0510	510
3	J4740-01	STORM-DRAINS-1	1.3844	1.3844	900	1.3852	1.3852	1.3852	0.0008	0.9
4	J4740-02	STORM-DRAINS-2	1.3843	1.3843	900	1.3849	1.3849	1.3849	0.0006	0.7
5	J4758-02	COMP	1.4017	1.4017	50	1.4033	1.4033	1.4033	0.0016	32
6	J4763-01	001M-WILLETS-PT-BLVD (AUG)	1.3944	1.3944	100	1.3965	1.3965	1.3965	0.0021	21
7	J4763-01DUP	001M-WILLETS-PT-BLVD (AUG) D	1.3946	1.3946	100	1.3966	1.3966	1.3966	0.0020	20
8	J4763-02	002M-35TH-AVE (AUG)	1.4028	1.4028	100	1.4107	1.4107	1.4107	0.0079	79

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS090418-2

WorkList ID : 116318

Date : 9/4/2018 7:29:46 AM

2397694

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	J4740-01	TSS	Cool 4 deg C	KWAR01	H33	STORM-DRAINS-1	08/30/2018	SM2540 D
	Water	J4740-02	TSS	Cool 4 deg C	KWAR01	H33	STORM-DRAINS-2	08/30/2018	SM2540 D
09/07/2018	Water	J4758-02	-	Cool 4 deg C	ARAM01	H52	COMP	08/31/2018	SM2540 D
09/06/2018	Water	J4763-01	TSS	Cool 4 deg C	TULL01	H51	001M-WILLETS-PT-BLVD(AU	08/30/2018	SM2540 D
09/06/2018	Water	J4763-02	TSS	Cool 4 deg C	TULL01	H51	002M-35TH-AVE(AUG)	08/30/2018	SM2540 D

Date/Time

Received by:

Relinquished by:

09/04/18, 08:00
AS
AS

Date/Time

Received by:

Relinquished by:

09/04/18, 12:30
AS
AS