


LB 97699
INSTRUMENT ID: CV2

7470A

Linear

μ Abs.:

38336



A= 0.0000e+000

B= 2.6479e-004

C= 4.8523e-003

Rho= 0.9989497

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.047	0.047	161	0.000	161				
0.2	0.200	0.195	-0.005	719	0.0 %	719				
2.5	2.500	2.460	-0.040	9271	0.0 %	9271				
5.0	5.000	5.177	0.177	19531	0.0 %	19531				
7.5	7.500	7.165	-0.335	27041	0.0 %	27041				
10.0	10.000	10.156	0.156	38336	0.0 %	38336				

LB97699 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 04 Sep 2018 11:18:49

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	30	161	-	0.0000	7470A	PPB	04 Sep 2018 11:20:12	S	Std
0.2 - 1	50.2	719	-	0.2000	7470A	PPB	04 Sep 2018 11:22:20	S	Std
2.5 - 1	52.5	9271	-	2.5000	7470A	PPB	04 Sep 2018 11:24:29	S	Std
5.0 - 1	55	19531	-	5.0000	7470A	PPB	04 Sep 2018 11:26:37	S	Std
7.5 - 1	59.5	27041	-	7.5000	7470A	PPB	04 Sep 2018 11:28:45	S	Std
10.0 - 1	516	38336	-	10.0000	7470A	PPB	04 Sep 2018 11:33:38	S	Std
ICV45 - 1	ICV45	15469	4.1009	-	7470A	PPB	04 Sep 2018 11:36:21	U	SMPL
ICB45 - 1	ICB45	-118	-0.0264	-	7470A	PPB	04 Sep 2018 11:38:27	U	SMPL
CCV63 - 1	CCV63	19289	5.1124	-	7470A	PPB	04 Sep 2018 11:40:34	U	SMPL
CCB63 - 1	CCB63	-102	-0.0222	-	7470A	PPB	04 Sep 2018 11:42:41	U	SMPL
CRA - 1	CRA	683	0.1857	-	7470A	PPB	04 Sep 2018 11:44:48	U	SMPL
HighStd - 1	HighStd	37481	9.9295	-	7470A	PPB	04 Sep 2018 11:46:54	U	SMPL
ChkStd - 1	ChkStd	25812	6.8397	-	7470A	PPB	04 Sep 2018 11:49:01	U	SMPL
PB112533BL - 1	PBW	-25	-0.0018	-	7470A	PPB	04 Sep 2018 11:54:22	U	SMPL
PB112533BS - 1	LCSW	13304	3.5276	-	7470A	PPB	04 Sep 2018 11:56:29	U	SMPL
PB112533TB - 1	PB1125515TB	-111	-0.0245	-	7470A	PPB	04 Sep 2018 11:58:36	U	SMPL
J4735-02 - 1	KG-BASELINE-WATER	-26	-0.0020	-	7470A	PPB	04 Sep 2018 12:00:44	U	SMPL
J4735-02DUP - 1	KG-BASELINE-WATERDUP	-18	0.0001	-	7470A	PPB	04 Sep 2018 12:02:52	U	SMPL
J4735-02MS - 1	KG-BASELINE-WATERMS	14515	3.8483	-	7470A	PPB	04 Sep 2018 12:05:02	U	SMPL
J4735-02MSD - 1	KG-BASELINE-WATERMSD	13865	3.6762	-	7470A	PPB	04 Sep 2018 12:07:08	U	SMPL
CCV64 - 1	CCV64	19101	5.0626	-	7470A	PPB	04 Sep 2018 12:09:16	U	SMPL
CCB64 - 1	CCB64	-112	-0.0248	-	7470A	PPB	04 Sep 2018 12:11:23	U	SMPL
PB112538BL - 1	PBW	-14	0.0011	-	7470A	PPB	04 Sep 2018 12:13:31	U	SMPL
PB112538BS - 1	LCSW	13622	3.6119	-	7470A	PPB	04 Sep 2018 12:15:37	U	SMPL
PB112538TB - 1	LB112509TB	-107	-0.0235	-	7470A	PPB	04 Sep 2018 12:17:44	U	SMPL
J4270-06 - 1	BU-02-083018	-2	0.0043	-	7470A	PPB	04 Sep 2018 12:19:51	U	SMPL
J4270-06DUP - 1	BU-02-083018DUP	-9	0.0025	-	7470A	PPB	04 Sep 2018 12:21:58	U	SMPL
J4270-06MS - 1	BU-02-083018MS	13711	3.6354	-	7470A	PPB	04 Sep 2018 12:24:06	U	SMPL
J4270-06MSD - 1	BU-02-083018MSD	13432	3.5615	-	7470A	PPB	04 Sep 2018 12:26:14	U	SMPL
J4720-06 - 1	BZTRANS-1	-118	-0.0264	-	7470A	PPB	04 Sep 2018 12:28:22	U	SMPL
J4720-07 - 1	BZTRANS-2	46	0.0170	-	7470A	PPB	04 Sep 2018 12:30:31	U	SMPL
J4720-08 - 1	BZTRANS-3	-13	0.0014	-	7470A	PPB	04 Sep 2018 12:32:37	U	SMPL
CCV65 - 1	CCV65	19231	5.0971	-	7470A	PPB	04 Sep 2018 12:34:45	U	SMPL
CCB65 - 1	CCB65	-51	-0.0087	-	7470A	PPB	04 Sep 2018 12:36:52	U	SMPL
J4720-09 - 1	BZTRANS-4	-10	0.0022	-	7470A	PPB	04 Sep 2018 12:39:00	U	SMPL
J4720-10 - 1	BZTRANS-5	20	0.0101	-	7470A	PPB	04 Sep 2018 12:41:07	U	SMPL
J4721-01 - 1	CB-0818-S-TCLP-GRID42(4)	6	0.0064	-	7470A	PPB	04 Sep 2018 12:43:16	U	SMPL
J4725-01 - 1	CONCRETE-BEDROCK	-18	0.0001	-	7470A	PPB	04 Sep 2018 12:45:23	U	SMPL
J4726-02 - 1	KG-SLUDGE-1	160	0.0472	-	7470A	PPB	04 Sep 2018 12:47:30	U	SMPL
J4726-04 - 1	KG-SLUDGE-2	194	0.0562	-	7470A	PPB	04 Sep 2018 12:49:37	U	SMPL
PB112539BL - 1	PBW	-24	-0.0015	-	7470A	PPB	04 Sep 2018 12:51:45	U	SMPL
PB112539BS - 1	LCSW	12768	3.3857	-	7470A	PPB	04 Sep 2018 12:53:53	U	SMPL
J4735-01 - 1	KG-BASELINE-WATER	-29	-0.0028	-	7470A	PPB	04 Sep 2018 12:56:02	U	SMPL
J4735-01DUP - 1	KG-BASELINE-WATERDUP	98	0.0308	-	7470A	PPB	04 Sep 2018 12:58:09	U	SMPL
CCV66 - 1	CCV66	19284	5.1111	-	7470A	PPB	04 Sep 2018 13:00:16	U	SMPL
CCB66 - 1	CCB66	-32	-0.0036	-	7470A	PPB	04 Sep 2018 13:02:24	U	SMPL
J4735-01MS - 1	KG-BASELINE-WATERMS	14160	3.7543	-	7470A	PPB	04 Sep 2018 13:04:33	U	SMPL
J4735-01MSD - 1	KG-BASELINE-WATERMSD	13890	3.6828	-	7470A	PPB	04 Sep 2018 13:06:40	U	SMPL
J4736-01 - 1	WC-1	-110	-0.0243	-	7470A	PPB	04 Sep 2018 13:08:48	U	SMPL
J4735-02LX5 - 1		49	0.0178	-	7470A	PPB	04 Sep 2018 13:10:56	U	SMPL
J4735-02A - 1		14422	3.8237	-	7470A	PPB	04 Sep 2018 13:13:03	U	SMPL
J4270-06LX5 - 1		-26	-0.0020	-	7470A	PPB	04 Sep 2018 13:15:11	U	SMPL
J4270-06A - 1		14329	3.7991	-	7470A	PPB	04 Sep 2018 13:17:20	U	SMPL
J4735-01LX5 - 1		-21	-0.0007	-	7470A	PPB	04 Sep 2018 13:19:29	U	SMPL
J4735-01A - 1		16043	4.2529	-	7470A	PPB	04 Sep 2018 13:21:38	U	SMPL
CCV67 - 1	CCV67	19185	5.0849	-	7470A	PPB	04 Sep 2018 13:23:45	U	SMPL
CCB67 - 1	CCB67	-91	-0.0192	-	7470A	PPB	04 Sep 2018 13:25:53	U	SMPL