



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB97714

Slope : 99.0

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.0	4.01	09/04/2018	16:00
2	CAL2	1	Water			21.9	7.01	09/04/2018	16:02
3	CAL3	1	Water			21.8	10.04	09/04/2018	16:04
4	ICV	1	Water			21.6	7.00	09/04/2018	16:06
5	CCV1	1	Water			21.8	2.00	09/04/2018	16:08
6	J4789-01	1	Solid	20.02	20	24.5	7.71	09/04/2018	16:10
7	J4789-01DUP	1	Solid	20.03	20	24.6	7.72	09/04/2018	16:12
8	J4789-02	1	Solid	20.03	20	24.1	8.21	09/04/2018	16:14
9	J4789-03	1	Solid	20.02	20	23.8	7.72	09/04/2018	16:16
10	CCV2	1	Water			21.8	12.00	09/04/2018	16:18

WORKLIST(Hardcopy Internal Chain)

LB97714

WorkList Name : pH090418

WorkList ID : 116373

Date : 9/4/2018 3:48:09 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/04/2018	Solid	J4789-01	pH	Cool 4 deg C	CTIA01	H42	WJ-18-09-04-TSHG-06	09/04/2018	9045D
09/04/2018	Solid	J4789-02	pH	Cool 4 deg C	CTIA01	H42	WJ-18-09-04-TSHG-07	09/04/2018	9045D
09/04/2018	Solid	J4789-03	pH	Cool 4 deg C	CTIA01	H42	WJ-18-09-04-TSHG-08	09/04/2018	9045D

Date/Time 09/04/18 3:55

Received by: SP

Relinquished by: CP

Date/Time 09/04/18 4:30

Received by: CP

Relinquished by: SP