


7470A

LB 97718
INSTRUMENT ID: CV4

Linear

 μ Abs.:

36555



A= 0.0000e+000

B= 2.7535e-004

C= 3.2216e-002

Rho= 0.9997941

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.024	0.024	-29	0.000	-29				
0.2	0.200	0.211	0.011	648	0.0 %	648				
2.5	2.500	2.490	-0.010	8925	0.0 %	8925				
5.0	5.000	5.028	0.028	18143	0.0 %	18143				
7.5	7.500	7.350	-0.150	26576	0.0 %	26576				
10.0	10.000	10.098	0.098	36555	0.0 %	36555				

LB97718 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 05 Sep 2018 08:26:55

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-29	-	0.0000	7470A	PPB	05 Sep 2018 08:30:53	S	Std
0.2 - 1	50.2	648	-	0.2000	7470A	PPB	05 Sep 2018 08:33:01	S	Std
2.5 - 1	52.5	8925	-	2.5000	7470A	PPB	05 Sep 2018 08:35:09	S	Std
5.0 - 1	55	18143	-	5.0000	7470A	PPB	05 Sep 2018 08:37:18	S	Std
7.5 - 1	57.5	26576	-	7.5000	7470A	PPB	05 Sep 2018 08:39:26	S	Std
10.0 - 1	510	36555	-	10.0000	7470A	PPB	05 Sep 2018 08:41:35	S	Std
ICV47 - 1	ICV47	14149	3.9281	-	7470A	PPB	05 Sep 2018 08:44:27	U	SMPL
ICB47 - 1	ICB47	-186	-0.0190	-	7470A	PPB	05 Sep 2018 08:46:34	U	SMPL
CCV75 - 1	CCV75	18198	5.0430	-	7470A	PPB	05 Sep 2018 08:48:41	U	SMPL
CCB75 - 1	CCB75	-136	-0.0052	-	7470A	PPB	05 Sep 2018 08:50:48	U	SMPL
CRA - 1	CRA	635	0.2071	-	7470A	PPB	05 Sep 2018 08:52:56	U	SMPL
HighStd - 1	HighStd	36542	10.0940	-	7470A	PPB	05 Sep 2018 08:55:02	U	SMPL
ChkStd - 1	ChkStd	25588	7.0779	-	7470A	PPB	05 Sep 2018 08:57:09	U	SMPL
PB112600BL - 1	PBW	-240	-0.0339	-	7470A	PPB	05 Sep 2018 08:59:17	U	SMPL
PB112600BS - 1	LCSW	13945	3.8720	-	7470A	PPB	05 Sep 2018 09:01:27	U	SMPL
PB112600TB - 1	PB112552TB	-65	0.0143	-	7470A	PPB	05 Sep 2018 09:03:35	U	SMPL
J4523-04 - 1	WC-2	48	0.0454	-	7470A	PPB	05 Sep 2018 09:05:44	U	SMPL
J4747-04 - 1	FL-DRUM-3-STONE	33	0.0413	-	7470A	PPB	05 Sep 2018 09:07:52	U	SMPL
J4756-03 - 1	WE-3	63	0.0496	-	7470A	PPB	05 Sep 2018 09:10:01	U	SMPL
J4756-04 - 1	WE-4	10	0.0350	-	7470A	PPB	05 Sep 2018 09:12:07	U	SMPL
CCV76 - 1	CCV76	18453	5.1132	-	7470A	PPB	05 Sep 2018 09:14:14	U	SMPL
CCB76 - 1	CCB76	-109	0.0022	-	7470A	PPB	05 Sep 2018 09:16:20	U	SMPL
J4757-02 - 1	WC-BR	-3	0.0314	-	7470A	PPB	05 Sep 2018 09:18:29	U	SMPL
J4760-03 - 1	LS-OILY-DEBRI	58	0.0482	-	7470A	PPB	05 Sep 2018 09:20:35	U	SMPL
J4761-03 - 1	RT-4005	-48	0.0190	-	7470A	PPB	05 Sep 2018 09:22:42	U	SMPL
J4761-06 - 1	RBR-200031	17	0.0369	-	7470A	PPB	05 Sep 2018 09:24:49	U	SMPL
J4764-02 - 1	SP	36	0.0421	-	7470A	PPB	05 Sep 2018 09:26:56	U	SMPL
J4764-02DUP - 1	SPDUP	88	0.0564	-	7470A	PPB	05 Sep 2018 09:29:04	U	SMPL
J4764-02MS - 1	SPMS	13764	3.8221	-	7470A	PPB	05 Sep 2018 09:31:12	U	SMPL
J4764-02MSD - 1	SPMSD	13130	3.6476	-	7470A	PPB	05 Sep 2018 09:33:20	U	SMPL
PB112606BL - 1	PBW	-99	0.0050	-	7470A	PPB	05 Sep 2018 09:35:29	U	SMPL
PB112606BS - 1	LCSW	14121	3.9204	-	7470A	PPB	05 Sep 2018 09:37:38	U	SMPL
CCV77 - 1	CCV77	18376	5.0920	-	7470A	PPB	05 Sep 2018 09:39:44	U	SMPL
CCB77 - 1	CCB77	-184	-0.0184	-	7470A	PPB	05 Sep 2018 09:41:52	U	SMPL
J4783-01 - 1	P001-GW001-01	1534	0.4546	-	7470A	PPB	05 Sep 2018 09:44:00	U	SMPL
J4783-01DUP - 1	P001-GW001-01DUP	1431	0.4262	-	7470A	PPB	05 Sep 2018 09:46:07	U	SMPL
J4783-01MS - 1	P001-GW001-01MS	18464	5.1163	-	7470A	PPB	05 Sep 2018 09:48:14	U	SMPL
J4783-01MSD - 1	P001-GW001-01MSD	18297	5.0703	-	7470A	PPB	05 Sep 2018 09:56:34	U	SMPL
J4764-02LX5 - 1		-80	0.0102	-	7470A	PPB	05 Sep 2018 10:08:32	U	SMPL
J4764-02A - 1		12608	3.5038	-	7470A	PPB	05 Sep 2018 10:10:40	U	SMPL
J4783-01LX5 - 1		-18	0.0273	-	7470A	PPB	05 Sep 2018 10:12:49	U	SMPL
J4783-01A - 1		18206	5.0452	-	7470A	PPB	05 Sep 2018 10:14:59	U	SMPL
CCV78 - 1	CCV78	18359	5.0874	-	7470A	PPB	05 Sep 2018 10:17:06	U	SMPL
CCB78 - 1	CCB78	-171	-0.0149	-	7470A	PPB	05 Sep 2018 10:19:16	U	SMPL