



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB97719

Slope : 99.0

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.0	4.01	09/04/2018	16:00
2	CAL2	1	Water			21.9	7.01	09/04/2018	16:02
3	CAL3	1	Water			21.8	10.04	09/04/2018	16:04
4	ICV	1	Water			21.6	7.00	09/04/2018	16:06
5	CCV1	1	Water			21.7	2.00	09/04/2018	16:35
6	J4781-02	1	Solid	20.02	20	24.1	8.94	09/04/2018	16:37
7	J4781-02DUP	1	Solid	20.03	20	24.2	8.95	09/04/2018	16:39
8	CCV2	1	Water			21.8	12.00	09/04/2018	16:41

WORKLIST(Hardcopy Internal Chain)

LB 97719

WorkList Name : corrosivity090418

WorkList ID : 116371

Date : 9/4/2018 2:43:56 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/04/2018	Solid	J4781-02	Corrosivity	Cool 4 deg C	PSEG01	I11	BS-STONE	09/04/2018	9045D

Date/Time 09/04/18 3:55
 Received by: SP
 Relinquished by: CP

Date/Time 09/04/18 4:50
 Received by: CP
 Relinquished by: SP