

LB97723


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

45264



A= 0.0000e+000

B= 2.2016e-004

C= -5.2974e-002

Rho= 0.9998016

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.058	-0.058	-25	0.000	-25				
0.2	0.200	0.225	0.025	1261	0.0 %	1261				
2.5	2.500	2.538	0.038	11769	0.0 %	11769				
5.0	5.000	4.953	-0.047	22736	0.0 %	22736				
7.5	7.500	7.631	0.131	34899	0.0 %	34899				
10.0	10.000	9.913	-0.087	45264	0.0 %	45264				

LB97723 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 05 Sep 2018 10:57:57

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc.	Method	Units	Date	Type	Type
0.0 - 1	30	-25	-	0.00007471B	PPB		05 Sep 2018 11:05:14	S	Std
0.2 - 1	50.2	1261	-	0.20007471B	PPB		05 Sep 2018 11:07:31	S	Std
2.5 - 1	52.5	11769	-	2.50007471B	PPB		05 Sep 2018 11:09:49	S	Std
5.0 - 1	55	22736	-	5.00007471B	PPB		05 Sep 2018 11:12:07	S	Std
7.5 - 1	57.5	34899	-	7.50007471B	PPB		05 Sep 2018 11:14:24	S	Std
10.0 - 1	510	45264	-	10.00007471B	PPB		05 Sep 2018 11:16:42	S	Std
ICV48 - 1	ICV48	18693	4.0625	-7471B	PPB		05 Sep 2018 11:20:14	U	SMPL
ICB48 - 1	ICB48	142	-0.0217	-7471B	PPB		05 Sep 2018 11:22:30	U	SMPL
CCV79 - 1	CCV79	22955	5.0009	-7471B	PPB		05 Sep 2018 11:24:48	U	SMPL
CCB79 - 1	CCB79	93	-0.0325	-7471B	PPB		05 Sep 2018 11:27:04	U	SMPL
CRA - 1	CRA	1160	0.2024	-7471B	PPB		05 Sep 2018 11:29:21	U	SMPL
HighStd - 1	HighStd	45767	10.0233	-7471B	PPB		05 Sep 2018 11:31:37	U	SMPL
ChkStd - 1	ChkStd	32548	7.1129	-7471B	PPB		05 Sep 2018 11:33:53	U	SMPL
PB112612BL - 1	PBS	13	-0.0501	-7471B	PPB		05 Sep 2018 11:36:10	U	SMPL
PB112612BS - 1	LCSS	16536	3.5877	-7471B	PPB		05 Sep 2018 11:38:30	U	SMPL
J4781-01 - 1	BS-STONES	305	0.0142	-7471B	PPB		05 Sep 2018 11:40:48	U	SMPL
J4782-02 - 1	ENY-SB-04-05-06-07-08-09-9.5-10	8591	1.8385	-7471B	PPB		05 Sep 2018 11:47:54	U	SMPL
J4782-02DUP - 1	ENY-SB-04-05-06-07-08-09-9.5-10DUP	8349	1.7852	-7471B	PPB		05 Sep 2018 11:50:11	U	SMPL
J4782-03 - 1	J4782-02MS	25309	5.5191	-7471B	PPB		05 Sep 2018 11:52:29	U	SMPL
J4782-04 - 1	J4782-02MSD	26563	5.7952	-7471B	PPB		05 Sep 2018 11:54:46	U	SMPL
CCV80 - 1	CCV80	22385	4.8754	-7471B	PPB		05 Sep 2018 11:57:02	U	SMPL
CCB80 - 1	CCB80	86	-0.0340	-7471B	PPB		05 Sep 2018 11:59:19	U	SMPL
J4782-05 - 1	ENY-SB-104-05-06-07-08-09-9.5-10	11494	2.4776	-7471B	PPB		05 Sep 2018 12:01:36	U	SMPL
J4789-01 - 1	WJ-18-09-04-TSHG-06	1905	0.3664	-7471B	PPB		05 Sep 2018 12:03:52	U	SMPL
J4789-02 - 1	WJ-18-09-04-TSHG-07	1887	0.3625	-7471B	PPB		05 Sep 2018 12:06:08	U	SMPL
J4789-03 - 1	WJ-18-09-04-TSHG-08	1764	0.3354	-7471B	PPB		05 Sep 2018 12:08:24	U	SMPL
J4791-01 - 1	DOCUMENTATION-VST	45109	9.8784	-7471B	PPB		05 Sep 2018 12:10:41	U	SMPL
J4791-02 - 1	DOCUMENTATION-#11	21247	4.6248	-7471B	PPB		05 Sep 2018 12:12:57	U	SMPL
J4791-03 - 1	DOCUMENTATION-#12	12222	2.6379	-7471B	PPB		05 Sep 2018 12:15:16	U	SMPL
J4791-04 - 1	DOCUMENTATION-#13	34169	7.4698	-7471B	PPB		05 Sep 2018 12:17:36	U	SMPL
J4791-05 - 1	DOCUMENTATION-#14	20657	4.4949	-7471B	PPB		05 Sep 2018 12:19:54	U	SMPL
J4791-06 - 1	DOCUMENTATION-#15	15578	3.3767	-7471B	PPB		05 Sep 2018 12:22:11	U	SMPL
CCV81 - 1	CCV81	22508	4.9025	-7471B	PPB		05 Sep 2018 12:24:29	U	SMPL
CCB81 - 1	CCB81	182	-0.0129	-7471B	PPB		05 Sep 2018 12:26:45	U	SMPL
J4791-07 - 1	DOCUMENTATION-#16	33989	7.4302	-7471B	PPB		05 Sep 2018 12:29:03	U	SMPL
J4791-08 - 1	DOCUMENTATION-#17	40499	8.8634	-7471B	PPB		05 Sep 2018 12:31:20	U	SMPL
J4782-02LX5 - 1		1983	0.3836	-7471B	PPB		05 Sep 2018 12:33:39	U	SMPL
J4782-02A - 1		28123	6.1387	-7471B	PPB		05 Sep 2018 12:35:57	U	SMPL
CCV82 - 1	CCV82	22356	4.8690	-7471B	PPB		05 Sep 2018 12:38:13	U	SMPL
CCB82 - 1	CCB82	14	-0.0499	-7471B	PPB		05 Sep 2018 12:40:32	U	SMPL