



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB97733

Slope : 98.4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.8	4.01	09/05/2018	16:20
2	CAL2	1	Water			21.9	7.01	09/05/2018	16:22
3	CAL3	1	Water			21.8	10.04	09/05/2018	16:24
4	ICV	1	Water			21.7	7.00	09/05/2018	16:26
5	CCV1	1	Water			21.9	2.00	09/05/2018	16:40
6	J4795-04	1	Water			20.1	5.26	09/05/2018	16:42
7	J4806-01	1	Water			20.2	6.35	09/05/2018	16:44
8	J4806-02	1	Water			20.7	6.63	09/05/2018	16:46
9	J4806-03	1	Water			20.7	6.72	09/05/2018	16:48
10	J4806-03DUP	1	Water			20.8	6.73	09/05/2018	16:50
11	CCV2	1	Water			21.8	12.00	09/05/2018	16:52

WORKLIST(Hardcopy Internal Chain)

LB 97733

WorkList Name : pH090518

WorkList ID : 116411

Date : 9/5/2018 2:17:00 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/05/2018	Water	J4795-04	pH	Cool 4 deg C	MAJO01	H61	DRAIN-WATER-TANK-1	09/05/2018	SM 4500-PH B
09/05/2018	Water	J4806-01	pH	Cool 4 deg C	ATCE02	H31	TAN-BLDG-1FL-KITCHEN-SIN	09/05/2018	SM 4500-PH B
09/05/2018	Water	J4806-02	pH	Cool 4 deg C	ATCE02	H31	BLDG-2FL-KITCHEN-SINK	09/05/2018	SM 4500-PH B
09/05/2018	Water	J4806-03	pH	Cool 4 deg C	ATCE02	H31	RED-BLDG-3FL-BATH-SINK	09/05/2018	SM 4500-PH B

Date/Time 09/05/18 4:15
 Received by: SP
 Relinquished by: CP

Date/Time 09/05/18 5:00
 Received by: CP
 Relinquished by: SP