



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB97735

Reagent/Standard	Lot/Log #
Residual chlorine std, 0.8PPM	WP67555
Residual chlorine std, 0PPM	WP67550
Residual chlorine std, 0.1PPM	WP67551
Residual chlorine std, 0.2PPM	WP67553
Residual chlorine std, and CCV , 0.4PPM	WP67554
Residual chlorine std, 1.6PPM	WP67556
Residual chlorine ICV-LCS, 0.4PPM	WP67557
Total Chlorine Powder Pillows	W1960

Intercept: -0.0057

Slope: 1.0240

Regression: 0.999922

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	09/05/2018	16:32
2	CAL2	0.1	1	0.000	0.100	0.100	09/05/2018	16:33
3	CAL3	0.2	1	0.000	0.200	0.200	09/05/2018	16:34
4	CAL4	0.4	1	0.000	0.400	0.400	09/05/2018	16:35
5	CAL5	0.8	1	0.000	0.800	0.800	09/05/2018	16:36
6	CAL6	1.6	1	0.000	1.640	1.640	09/05/2018	16:37



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB97735

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.000	0.410	0.410	0.406	09/05/2018	16:38
2	ICB				1	0.000	0.000	0.000	0.006	09/05/2018	16:39
3	CCV1			0.4	1	0.000	0.380	0.380	0.377	09/05/2018	16:40
4	CCB1				1	0.000	0.000	0.000	0.006	09/05/2018	16:41
5	LB97735BL	50	50		1	0.000	0.000	0.000	0.006	09/05/2018	16:42
6	LB97735BS	50	50	0.4	1	0.000	0.410	0.410	0.406	09/05/2018	16:43
7	J4806-01	50	50		1	0.000	0.030	0.030	0.035	09/05/2018	16:44
8	J4806-02	50	50		1	0.010	0.100	0.090	0.093	09/05/2018	16:45
9	J4806-03	50	50		1	0.010	0.130	0.120	0.123	09/05/2018	16:46
10	J4806-03DUP	50	50		1	0.000	0.120	0.120	0.123	09/05/2018	16:47
11	J4806-03MS	50	50	0.4	1	0.000	0.530	0.530	0.523	09/05/2018	16:49
12	J4806-03MSD	50	50	0.4	1	0.010	0.540	0.530	0.523	09/05/2018	16:50
13	CCV2			0.4	1	0.000	0.400	0.400	0.396	09/05/2018	16:51
14	CCB2				1	0.000	0.000	0.000	0.006	09/05/2018	16:52

LB97735

WORKLIST(Hardcopy Internal Chain)

WorkList Name : RS090518

WorkList ID : 116415

Date : 9/5/2018 3:47:24 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/05/2018	G Water	J4806-01	Residual Chlorine	Cool 4 deg C	ATCE02	H31	TAN-BLDG-1FL-KITCHEN-SIN	09/05/2018	SM4500 CI G
09/05/2018	G Water	J4806-02	Residual Chlorine	Cool 4 deg C	ATCE02	H31	RED-BLDG-2FL-KITCHEN-SII	09/05/2018	SM4500 CI G
09/05/2018	C Water	J4806-03	Residual Chlorine	Cool 4 deg C	ATCE02	H31	RED-BLDG-3FL-BATH-SINK	09/05/2018	SM4500 CI G

Date/Time 9-5-18 16:10
 Received by: AR
 Relinquished by: CP

Date/Time 9-5-18 17:00
 Received by: AR
 Relinquished by: AR