



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB97736

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.6	09/06/2018	09:35
2	J4801-06		1	8	>100.0	>212.0	>212.0	09/06/2018	09:50
3	J4801-06DUP		1	8	>100.0	>212.0	>212.0	09/06/2018	10:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 97736

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp9518

WorkList ID : 116422

Date : 9/5/2018 4:47:14 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/10/2018	Water	J4801-06	Flash Point	Cool 4 deg C	PSEG01	I13	OILY-WATER	09/05/2018	1010A

Date/Time 09/06/18 9:05 AM
 Received by: KS
 Relinquished by: CP

Date/Time 09/06/18 10:25 AM
 Received by: CP
 Relinquished by: KS