


ISM02.4

LB97742  
INSTRUMENT ID: CV1

Linear

 $\mu$  Abs.:

43330



A= 0.0000e+000

B= 2.3130e-004

C= 5.2943e-002

Rho= 0.9998853

Accept=Accepted

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|
| 0.00   | 0.000  | 0.043  | 0.043  | -42   | 0.000      | -42   |       |       |       |       |
| 0.2    | 0.200  | 0.199  | -0.001 | 632   | 0.0 %      | 632   |       |       |       |       |
| 2.5    | 2.500  | 2.472  | -0.028 | 10458 | 0.0 %      | 10458 |       |       |       |       |
| 5.0    | 5.000  | 5.012  | 0.012  | 21439 | 0.0 %      | 21439 |       |       |       |       |
| 7.5    | 7.500  | 7.399  | -0.101 | 31758 | 0.0 %      | 31758 |       |       |       |       |
| 10.0   | 10.000 | 10.075 | 0.075  | 43330 | 0.0 %      | 43330 |       |       |       |       |

## INSTRUMENT ID : CV1

| Sample ID  | Extended ID | $\mu$ Abs. | Conc.    | Std Conc | Method      | Units | Date          | Type |
|------------|-------------|------------|----------|----------|-------------|-------|---------------|------|
|            | 0 S0.0      | -42        | -        |          | 0 ISM02.4   | PPB   | 9/6/2018 7:57 | Std  |
|            | 0.2 S0.2    | 632        | -        |          | 0.2 ISM02.4 | PPB   | 9/6/2018 8:00 | Std  |
|            | 2.5 S2.5    | 10458      | -        |          | 2.5 ISM02.4 | PPB   | 9/6/2018 8:02 | Std  |
|            | 5 S5        | 21439      | -        |          | 5 ISM02.4   | PPB   | 9/6/2018 8:04 | Std  |
|            | 7.5 S7.5    | 31758      | -        |          | 7.5 ISM02.4 | PPB   | 9/6/2018 8:07 | Std  |
|            | 10 S10.0    | 43330      | -        |          | 10 ISM02.4  | PPB   | 9/6/2018 8:09 | Std  |
| ICV        | ICV         | 16645      | 3.903 -  |          | ISM02.4     | PPB   | 9/6/2018 8:12 | SMPL |
| ICB        | ICB         | -216       | 0.003 -  |          | ISM02.4     | PPB   | 9/6/2018 8:14 | SMPL |
| CCV085     | CCV085      | 21978      | 5.1365 - |          | ISM02.4     | PPB   | 9/6/2018 8:16 | SMPL |
| CCB085     | CCB085      | -175       | 0.0125 - |          | ISM02.4     | PPB   | 9/6/2018 8:19 | SMPL |
| PB112634BL | PBW010      | -92        | 0.0317 - |          | ISM02.4     | PPB   | 9/6/2018 8:21 | SMPL |
| PB112634TB | LEB010      | 71         | 0.0694 - |          | ISM02.4     | PPB   | 9/6/2018 8:23 | SMPL |
| J4766-04   | MCOAE3      | 74         | 0.0701 - |          | ISM02.4     | PPB   | 9/6/2018 8:25 | SMPL |
| J4766-05   | MCOAE3D     | 43         | 0.0629 - |          | ISM02.4     | PPB   | 9/6/2018 8:28 | SMPL |
| J4766-06   | MCOAE3S     | 4209       | 1.0265 - |          | ISM02.4     | PPB   | 9/6/2018 8:30 | SMPL |
| CCV086     | CCV086      | 21867      | 5.1108 - |          | ISM02.4     | PPB   | 9/6/2018 8:32 | SMPL |
| CCB086     | CCB086      | -73        | 0.0361 - |          | ISM02.4     | PPB   | 9/6/2018 8:34 | SMPL |