


ISM02.4

LB 97755
INSTRUMENT ID: CVD

Linear

 μ Abs.:

45804



A= 0.0000e+000

B= 2.1907e-004

C= -7.3311e-003

Rho= 0.9999461

Accept=Accepted

Slope
 $y = 2.1907e-004x - 7.3311e-003$

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9/10
0.00	0.000	0.021	0.021	128	0.000	128					-9
0.2	0.200	0.182	-0.018	863	0.0 %	863					-1
2.5	2.500	2.485	-0.015	11376	0.0 %	11376					-1
5.0	5.000	5.052	0.052	23095	0.0 %	23095					-1
7.5	7.500	7.434	-0.066	33967	0.0 %	33967					-1
10.0	10.000	10.027	0.027	45804	0.0 %	45804					0

ASB
9/16/18

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	128	-		0 ISM02.4	PPB	9/6/2018 13:40	Std
	0.2 S0.2	863	-		0.2 ISM02.4	PPB	9/6/2018 13:43	Std
	2.5 S2.5	11376	-		2.5 ISM02.4	PPB	9/6/2018 13:45	Std
	5 S5	23095	-		5 ISM02.4	PPB	9/6/2018 13:47	Std
	7.5 S7.5	33967	-		7.5 ISM02.4	PPB	9/6/2018 13:49	Std
	10 S10.0	45804	-		10 ISM02.4	PPB	9/6/2018 13:52	Std
ICV	ICV	17514	3.8294 -		ISM02.4	PPB	9/6/2018 13:55	SMPL
ICB	ICB	-54	-0.0192 -		ISM02.4	PPB	9/6/2018 13:57	SMPL
CCV087	CCV087	22957	5.0218 -		ISM02.4	PPB	9/6/2018 13:59	SMPL
CCB087	CCB087	-97	-0.0286 -		ISM02.4	PPB	9/6/2018 14:01	SMPL
PB112636BL	pbs004	-22	-0.0122 -		ISM02.4	PPB	9/6/2018 14:04	SMPL
J4666-01	ME0AA0	1415	0.3027 -		ISM02.4	PPB	9/6/2018 14:06	SMPL
J4666-02	ME0AA0D	1190	0.2534 -		ISM02.4	PPB	9/6/2018 14:08	SMPL
J4666-03	ME0AA0S	13018	2.8445 -		ISM02.4	PPB	9/6/2018 14:10	SMPL
J4666-04	ME0AA1	969	0.2049 -		ISM02.4	PPB	9/6/2018 14:13	SMPL
J4666-05	ME0AA2	8865	1.9347 -		ISM02.4	PPB	9/6/2018 14:15	SMPL
J4666-06	ME0AB4	8448	1.8434 -		ISM02.4	PPB	9/6/2018 14:17	SMPL
J4666-07	ME0AB7	1294	0.2761 -		ISM02.4	PPB	9/6/2018 14:20	SMPL
CCV088	CCV088	23126	5.0589 -		ISM02.4	PPB	9/6/2018 14:22	SMPL
CCB088	CCB088	-84	-0.0257 -		ISM02.4	PPB	9/6/2018 14:24	SMPL