

Analytical Summary Report

Analysis Method: SM5540 C

ANALYST: AMANDEEP

Parameter: MBAS

SUPERVISOR REVIEW BY: apatel

Run Number: LB97757

Reagent/Standard	Lot/Loq #
MBAS CAL STD, 0.5PPM	WP67580
MBAS CCV STD, 1.0PPM	WP67586
MBAS ICV-LCS std 1.0PPM	WP67587
MBAS DAILY CHECK STD, 0.01PPM	WP67588
MBAS DAILY CHECK STD, 2.0PPM	WP67589
MBAS CAL STD 2 PPM	WP67579
MBAS CAL STD 1 PPM	WP67578
MBAS CAL STD 0.1 PPM	WP67583
mbas cal std 0.05 ppm	WP67584
MBAS CAL STD 0.01 PPM	WP67585
MBAS BLANK 0 PPM	WP67577
MBAS CAL STD 0.4 PPM	WP67581
MBAS CAL STD 0.2 PPM	WP67582
Methylene blue reagent-1000ml	WP66778
MBAS Wash Solution	WP66800

Intercept: 0.0008

Slope: 0.2214

Regression: 0.9999

Seq	Lab ID	True Value (mg/L)	Dilution	Reading (652 nm)	Anal Date	Anal Time
1	CAL1	0	1	0.000	09/06/2018	12:10
2	CAL2	0.01	1	0.003	09/06/2018	12:10
3	CAL3	0.05	1	0.011	09/06/2018	12:11
4	CAL4	0.1	1	0.029	09/06/2018	12:11
5	CAL5	0.2	1	0.043	09/06/2018	12:12
6	CAL6	0.4	1	0.088	09/06/2018	12:13
7	CAL7	0.5	1	0.116	09/06/2018	12:13
8	CAL8	1	1	0.224	09/06/2018	12:14
9	CAL9	2	1	0.442	09/06/2018	12:15



Analytical Summary Report

Analysis Method: SM5540 C

ANALYST: AMANDEEP

Parameter: MBAS

SUPERVISOR REVIEW BY: apatel

Run Number: LB97757

Seq	Lab ID	True Value (mg/L)	Dilution	Reading (652 nm)	Result (mg/L)	AnalDate	Anal Time
1	ICV	1	1	0.211	0.949	09/06/2018	12:15
2	ICB		1	0.000	-0.004	09/06/2018	12:16
3	CHK1	0.01	1	0.003	0.010	09/06/2018	12:16
4	CHK2	2	1	0.409	1.844	09/06/2018	12:16
5	CCV1	1	1	0.217	0.977	09/06/2018	12:17
6	CCB1		1	0.000	-0.004	09/06/2018	12:17
7	LB97757BL		1	0.000	-0.004	09/06/2018	12:18
8	LB97757BS	1	1	0.209	0.940	09/06/2018	12:18
9	J4806-01		1	0.005	0.019	09/06/2018	12:19
10	J4806-02		1	0.004	0.014	09/06/2018	12:19
11	J4806-03		1	0.006	0.023	09/06/2018	12:20
12	J4806-03DUP		1	0.006	0.023	09/06/2018	12:20
13	J4806-03MS	0.5	1	0.123	0.552	09/06/2018	12:21
14	J4806-03MSD	0.5	1	0.122	0.547	09/06/2018	12:21
15	CCV2	1	1	0.219	0.986	09/06/2018	12:22
16	CCB2		1	0.000	-0.004	09/06/2018	12:23

WORKLIST(Hardcopy Internal Chain)

0907157

WorkList Name : mbas090618

WorkList ID : 116458

Date : 9/6/2018 10:30:53 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/07/2018	Water	J4806-01	MBAS	Cool 4 deg C	ATCE02	H31	TAN-BLDG-1FL-KITCHEN-SIN	09/05/2018	SM5540 C
09/07/2018	Water	J4806-02	MBAS	Cool 4 deg C	ATCE02	H31	RED-BLDG-2FL-KITCHEN-SII	09/05/2018	SM5540 C
09/07/2018	Water	J4806-03	MBAS	Cool 4 deg C	ATCE02	H31	RED-BLDG-3FL-BATH-SINK	09/05/2018	SM5540 C

Date/Time 9-6-18 10:00
 Received by: AP
 Relinquished by: CP

Date/Time 9-6-18 13:30
 Received by: CP
 Relinquished by: AP