

LB 97762


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

44393



A= 0.0000e+000

B= 2.2556e-004

C= 2.5812e-002

Rho= 0.9999454

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.029	0.029	14	0.000	14				
0.2	0.200	0.239	0.039	943	0.0 %	943				
2.5	2.500	2.438	-0.062	10695	0.0 %	10695				
5.0	5.000	4.967	-0.033	21906	0.0 %	21906				
7.5	7.500	7.488	-0.012	33082	0.0 %	33082				
10.0	10.000	10.039	0.039	44393	0.0 %	44393				

LB97762 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 06 Sep 2018 15:03:10

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	14	-	0.00007471B	PPB		06 Sep 2018 15:06:15	S	Std
0.2 - 1	50.2	943	-	0.20007471B	PPB		06 Sep 2018 15:08:33	S	Std
2.5 - 1	52.5	10695	-	2.50007471B	PPB		06 Sep 2018 15:10:51	S	Std
5.0 - 1	55	21906	-	5.00007471B	PPB		06 Sep 2018 15:13:08	S	Std
7.5 - 1	57.5	33082	-	7.50007471B	PPB		06 Sep 2018 15:18:07	S	Std
10.0 - 1	510	44393	-	10.00007471B	PPB		06 Sep 2018 15:25:35	S	Std
ICV49 - 1	ICV49	17798	4.0404	-7471B	PPB		06 Sep 2018 15:28:17	U	SMPL
ICB49 - 1	ICB49	-53	0.0139	-7471B	PPB		06 Sep 2018 15:30:35	U	SMPL
CCV89 - 1	CCV89	22290	5.0536	-7471B	PPB		06 Sep 2018 15:32:53	U	SMPL
CCB89 - 1	CCB89	-92	0.0051	-7471B	PPB		06 Sep 2018 15:35:09	U	SMPL
CRA - 1	CRA	906	0.2302	-7471B	PPB		06 Sep 2018 15:37:27	U	SMPL
HighStd - 1	HighStd	44169	9.9888	-7471B	PPB		06 Sep 2018 15:39:43	U	SMPL
ChkStd - 1	ChkStd	31023	7.0235	-7471B	PPB		06 Sep 2018 15:41:59	U	SMPL
PB112637BL - 1	PBS	-224	-0.0247	-7471B	PPB		06 Sep 2018 15:44:17	U	SMPL
PB112637BS - 1	LCSS	14947	3.3973	-7471B	PPB		06 Sep 2018 15:46:36	U	SMPL
J4801-01 - 1	CHRT-26957	1320	0.3236	-7471B	PPB		06 Sep 2018 15:48:53	U	SMPL
J4803-01 - 1	EP1-Q2-E5	1082	0.2699	-7471B	PPB		06 Sep 2018 15:51:10	U	SMPL
J4803-02 - 1	EP1-Q2-E6	841	0.2155	-7471B	PPB		06 Sep 2018 15:53:28	U	SMPL
J4803-03 - 1	EP1-Q2-E7	4931	1.1381	-7471B	PPB		06 Sep 2018 15:55:45	U	SMPL
J4803-04 - 1	EP1-Q2-F5	932	0.2360	-7471B	PPB		06 Sep 2018 15:58:02	U	SMPL
CCV90 - 1	CCV90	21838	4.9517	-7471B	PPB		06 Sep 2018 16:00:19	U	SMPL
CCB90 - 1	CCB90	-201	-0.0195	-7471B	PPB		06 Sep 2018 16:02:35	U	SMPL
J4803-05 - 1	EP1-Q2-F6	817	0.2101	-7471B	PPB		06 Sep 2018 16:04:53	U	SMPL
J4803-06 - 1	EP1-Q2-F7	790	0.2040	-7471B	PPB		06 Sep 2018 16:07:09	U	SMPL
J4803-07 - 1	EP1-Q2-G5	1145	0.2841	-7471B	PPB		06 Sep 2018 16:09:25	U	SMPL
J4803-08 - 1	EP1-Q2-G6	1263	0.3107	-7471B	PPB		06 Sep 2018 16:11:41	U	SMPL
J4803-09 - 1	EP1-Q2-G7	427	0.1221	-7471B	PPB		06 Sep 2018 16:13:57	U	SMPL
J4803-10 - 1	EP1-Q2-H5	1984	0.4733	-7471B	PPB		06 Sep 2018 16:16:15	U	SMPL
J4803-11 - 1	EP1-Q2-H6	892	0.2270	-7471B	PPB		06 Sep 2018 16:18:32	U	SMPL
J4803-12 - 1	EP1-Q2-H7	1247	0.3071	-7471B	PPB		06 Sep 2018 16:20:49	U	SMPL
J4803-13 - 1	EP1-Q2-I5	1746	0.4196	-7471B	PPB		06 Sep 2018 16:23:07	U	SMPL
J4803-14 - 1	EP1-Q2-I6	781	0.2020	-7471B	PPB		06 Sep 2018 16:25:24	U	SMPL
CCV91 - 1	CCV91	22148	5.0216	-7471B	PPB		06 Sep 2018 16:27:40	U	SMPL
CCB91 - 1	CCB91	-111	0.0008	-7471B	PPB		06 Sep 2018 16:29:56	U	SMPL
J4803-15 - 1	EP1-Q2-I7	987	0.2484	-7471B	PPB		06 Sep 2018 16:32:13	U	SMPL
J4803-16 - 1	EP1-Q2-J5	718	0.1878	-7471B	PPB		06 Sep 2018 16:34:29	U	SMPL
J4803-17 - 1	EP1-Q2-J6	1339	0.3278	-7471B	PPB		06 Sep 2018 16:36:48	U	SMPL
J4803-17DUP - 1	EP1-Q2-J6DUP	1521	0.3689	-7471B	PPB		06 Sep 2018 16:42:06	U	SMPL
J4803-18 - 1	EP1-Q2-J6-MS	18299	4.1534	-7471B	PPB		06 Sep 2018 16:44:23	U	SMPL
J4803-19 - 1	EP1-Q2-J6-MSD	17770	4.0341	-7471B	PPB		06 Sep 2018 16:46:40	U	SMPL
J4803-20 - 1	EP1-Q2-J7	1017	0.2552	-7471B	PPB		06 Sep 2018 16:48:58	U	SMPL
J4803-21 - 1	EP1-Q2-FD-01	530	0.1454	-7471B	PPB		06 Sep 2018 16:51:15	U	SMPL
J4803-17LX5 - 1		248	0.0818	-7471B	PPB		06 Sep 2018 16:53:34	U	SMPL
J4803-17A - 1		20381	4.6230	-7471B	PPB		06 Sep 2018 16:55:50	U	SMPL
CCV92 - 1	CCV92	22342	5.0654	-7471B	PPB		06 Sep 2018 16:58:07	U	SMPL
CCB92 - 1	CCB92	-79	0.0080	-7471B	PPB		06 Sep 2018 17:00:23	U	SMPL