



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB97763

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.0	4.01	09/06/2018	12:15
2	CAL2	1	Water			21.9	7.01	09/06/2018	12:17
3	CAL3	1	Water			22.0	10.04	09/06/2018	12:19
4	ICV	1	Water			22.2	7.00	09/06/2018	12:21
5	CCV1	1	Water			21.8	2.00	09/06/2018	12:23
6	J4808-01	1	Water			24.1	4.66	09/06/2018	12:25
7	J4808-01DUP	1	Water			24.2	4.67	09/06/2018	12:27
8	J4808-02	1	Water			23.5	4.60	09/06/2018	12:29
9	J4808-03	1	Water			23.6	4.66	09/06/2018	12:31
10	J4810-01	1	Water			24.3	4.13	09/06/2018	12:33
11	J4810-02	1	Water			23.9	8.04	09/06/2018	12:35
12	CCV2	1	Water			21.9	12.00	09/06/2018	12:37

WORKLIST(Hardcopy Internal Chain)

LB 97763

WorkList Name : pH090618

WorkList ID : 116461

Date : 9/6/2018 11:57:11 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/05/2018	Water	J4808-01	pH	Cool 4 deg C	STER13	H43	CELL-A-POOL-WATER	09/05/2018	SM 4500-PH B
09/05/2018	Water	J4808-02	pH	Cool 4 deg C	STER13	H43	CELL-B-POOL-WATER	09/05/2018	SM 4500-PH B
09/05/2018	Water	J4808-03	pH	Cool 4 deg C	STER13	H43	CELL-C-POOL-WATER	09/05/2018	SM 4500-PH B
09/05/2018	Water	J4810-01	pH	Cool 4 deg C	STER05	I12	POOL-WATER-DI-WATER	09/05/2018	SM 4500-PH B
09/05/2018	Water	J4810-02	pH	Cool 4 deg C	STER05	I12	CITY-WATER	09/05/2018	SM 4500-PH B

Date/Time 09/06/18 12:00

Received by: SP

Relinquished by: CP

Date/Time 09/06/18 1:00

Received by: CP

Relinquished by: SP