



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatelANALYST: AparnaDate: 09/06/2018TEMP1 IN: 103 09/06/2018 14:00 TEMP OUT: 104 09/06/2018 15:15TEMP2 IN: 104 09/06/2018 16:00 TEMP OUT: 104 09/07/2018 07:00TEMP3 IN: 105 09/07/2018 08:00 TEMP OUT: 105 09/07/2018 09:35TEMP4 IN: 105 09/07/2018 10:26 TEMP OUT: 105 09/07/2018 12:00Run Number: LB97769BalanceID: WC SC-3OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB97769BL	LB97769BL	1.3863	1.3863	100	1.3863	1.3863	1.3863	0.0000	0
2	LB97769BS	LB97769BS	1.3794	1.3794	100	1.4296	1.4296	1.4296	0.0502	502
3	J4813-01	DRAIN-WATER-TANK-1	1.3933	1.3933	20	1.4057	1.4057	1.4057	0.0124	620
4	J4814-01	32220008-3	1.3821	1.3821	100	1.3867	1.3867	1.3867	0.0046	46
5	J4815-01	SPDES-1	1.3857	1.3857	900	1.3871	1.3871	1.3871	0.0014	1.6
6	J4819-01	001M-WILLETS-PT-BLVD (SEP)	1.3932	1.3932	50	1.3988	1.3988	1.3988	0.0056	112
7	J4819-02	002M-35TH-AVE (SEP)	1.3903	1.3903	100	1.3942	1.3942	1.3942	0.0039	39
8	J4819-02DUP	002M-35TH-AVE (SEP) DUP	1.4073	1.4073	100	1.4113	1.4113	1.4113	0.0040	40

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS090718-2

WorkList ID : 116471

Date : 9/7/2018 7:33:18 AM

21397769

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/13/2018	Water	J4813-01	TSS	Cool 4 deg C	MAJO01	I21	DRAIN-WATER-TANK-1	09/06/2018	SM2540 D
09/13/2018	Water	J4814-01	TSS	Cool 4 deg C	PHAR02	I51	32220008-3	09/06/2018	SM2540 D
09/07/2018	Water	J4815-01	TSS	Cool 4 deg C	SKAN06	I23	SPDES-1	09/06/2018	SM2540 D
09/10/2018	Water	J4819-01	TSS	Cool 4 deg C	TULL01	I23	001M-WILLETS-PT-BLVD(SEI	09/05/2018	SM2540 D
09/10/2018	Water	J4819-02	TSS	Cool 4 deg C	TULL01	I23	002M-35TH-AVE(SEP)	09/05/2018	SM2540 D

Date/Time

Received by:

Relinquished by:

09/07/18, 04:40
AS
CS

Date/Time

Received by:

Relinquished by:

09/07/18, 08:46
AS
AS