



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB97774

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.8	4.01	09/06/2018	17:10
2	CAL2	1	Water			22.0	7.01	09/06/2018	17:12
3	CAL3	1	Water			21.8	10.04	09/06/2018	17:14
4	ICV	1	Water			21.9	7.00	09/06/2018	17:16
5	CCV1	1	Water			21.9	2.00	09/06/2018	17:18
6	J4774-02	1	Solid	20.02	20	25.9	8.62	09/06/2018	17:20
7	J4774-02DUP	1	Solid	20.03	20	26.0	8.63	09/06/2018	17:22
8	J4775-02	1	Solid	20.02	20	25.8	8.32	09/06/2018	17:24
9	J4775-04	1	Solid	20.03	20	25.8	8.46	09/06/2018	17:26
10	J4779-02	1	Solid	20.03	20	25.6	8.48	09/06/2018	17:28
11	J4779-04	1	Solid	20.04	20	25.7	8.30	09/06/2018	17:30
12	J4809-04	1	Solid	20.02	20	25.6	8.80	09/06/2018	17:32
13	J4809-08	1	Solid	20.03	20	26.0	7.12	09/06/2018	17:34
14	J4825-19	1	Solid	20.02	20	25.5	8.34	09/06/2018	17:36
15	J4825-20	1	Solid	20.02	20	25.7	8.22	09/06/2018	17:38
16	CCV2	1	Water			21.8	12.00	09/06/2018	17:40

WORKLIST(Hardcopy Internal Chain)

LB 97774

WorkList Name : corrosivity090618

WorkList ID : 116468

Date : 9/6/2018 4:30:53 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/06/2018	Solid	J4825-19	Corrosivity	Cool 4 deg C	CHAZ02		EP1-WS-COMP-17	09/06/2018	9045D
09/06/2018	Solid	J4825-20	Corrosivity	Cool 4 deg C	CHAZ02		EP1-WS-COMP-18	09/06/2018	9045D
09/06/2018	Solid	J4809-04	Corrosivity	Cool 4 deg C	PSEG01	I21	MS-DRUM-1-STONE	09/06/2018	9045D
09/06/2018	Solid	J4809-08	Corrosivity	Cool 4 deg C	PSEG01	I21	MS-SLUDGE-V24129	09/06/2018	9045D
09/06/2018	Solid	J4774-02	Corrosivity	Cool 4 deg C	PSEG05		OK-03-090618	09/06/2018	9045D
09/06/2018	Solid	J4775-02	Corrosivity	Cool 4 deg C	PSEG05	I12	OR-02-090618-A	09/06/2018	9045D
09/06/2018	Solid	J4775-04	Corrosivity	Cool 4 deg C	PSEG05	I12	OR-02-090618-B	09/06/2018	9045D
09/06/2018	Solid	J4779-02	Corrosivity	Cool 4 deg C	PSEG05	I21	SU-03-090618-A	09/06/2018	9045D
09/06/2018	Solid	J4779-04	Corrosivity	Cool 4 deg C	PSEG05	I21	SU-03-090618-B	09/06/2018	9045D

Date/Time 09/06/18 5:00

Received by: SP

Relinquished by: CP

Date/Time 09/06/18 5:50

Received by: CP

Relinquished by: SP