


LB 97780

245.1

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.8961e-004

C= 1.1848e-002

Rho= 0.9999270

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.027	0.027	53	0.000	53				
0.05	0.050	0.065	0.015	182	0.0 %	182				
0.2	0.200	0.192	-0.008	623	0.0 %	623				
2.5	2.500	2.419	-0.081	8311	0.0 %	8311				
5.0	5.000	5.024	0.024	17307	0.0 %	17307				
7.5	7.500	7.565	0.065	26079	0.0 %	26079				
10.0	10.000	9.958	-0.042	34345	0.0 %	34345				

LB97780 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 07 Sep 2018 10:09:38

Sample ID	Extended ID	µ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	53	-	0.0000245.1	PPB	07 Sep 2018 10:12:57	S	Std
0.05 - 1	50.05	182	-	0.0500245.1	PPB	07 Sep 2018 10:15:14	S	Std
0.2 - 1	50.2	623	-	0.2000245.1	PPB	07 Sep 2018 10:17:32	S	Std
2.5 - 1	52.5	8311	-	2.5000245.1	PPB	07 Sep 2018 10:19:49	S	Std
5.0 - 1	55	17307	-	5.0000245.1	PPB	07 Sep 2018 10:22:07	S	Std
7.5 - 1	57.5	26079	-	7.5000245.1	PPB	07 Sep 2018 10:24:23	S	Std
10.0 - 1	510	34345	-	10.0000245.1	PPB	07 Sep 2018 10:26:41	S	Std
ICV51 - 1	ICV51	13668	3.9702	-245.1	PPB	07 Sep 2018 10:30:07	U	SMPL
ICB51 - 1	ICB51	-275	-0.0678	-245.1	PPB	07 Sep 2018 10:32:23	U	SMPL
CCV01 - 1	CCV01	17124	4.9711	-245.1	PPB	07 Sep 2018 10:34:40	U	SMPL
CCB01 - 1	CCB01	-219	-0.0516	-245.1	PPB	07 Sep 2018 10:36:56	U	SMPL
CRA - 1	CRA	122	0.0472	-245.1	PPB	07 Sep 2018 10:39:14	U	SMPL
HighStd - 1	HighStd	35438	10.2750	-245.1	PPB	07 Sep 2018 10:41:30	U	SMPL
ChkStd - 1	ChkStd	24005	6.9639	-245.1	PPB	07 Sep 2018 10:46:32	U	SMPL
PB112650BL - 1	PBW	-317	-0.0800	-245.1	PPB	07 Sep 2018 10:48:48	U	SMPL
PB112650BS - 1	LCSW	12885	3.7435	-245.1	PPB	07 Sep 2018 10:51:07	U	SMPL
J4796-02 - 1	COMP	-154	-0.0328	-245.1	PPB	07 Sep 2018 10:53:23	U	SMPL
J4799-02 - 1	EFF-WASTE-WATER	65	0.0307	-245.1	PPB	07 Sep 2018 10:55:41	U	SMPL
J4806-01 - 1	TAN-BLDG-1FL-KITCHEN-SINK	-12	0.0084	-245.1	PPB	07 Sep 2018 10:57:59	U	SMPL
J4806-01DUP - 1	TAN-BLDG-1FL-KITCHEN-SINKDUP	-75	-0.0099	-245.1	PPB	07 Sep 2018 11:00:17	U	SMPL
J4806-01MS - 1	TAN-BLDG-1FL-KITCHEN-SINKMS	13292	3.8613	-245.1	PPB	07 Sep 2018 11:02:32	U	SMPL
CCV02 - 1	CCV02	17025	4.9424	-245.1	PPB	07 Sep 2018 11:04:48	U	SMPL
CCB02 - 1	CCB02	-272	-0.0669	-245.1	PPB	07 Sep 2018 11:07:05	U	SMPL
J4806-01MSD - 1	TAN-BLDG-1FL-KITCHEN-SINKMSD	13508	3.9239	-245.1	PPB	07 Sep 2018 11:09:23	U	SMPL
J4806-02 - 1	RED-BLDG-2FL-KITCHEN-SINK	-225	-0.0533	-245.1	PPB	07 Sep 2018 11:11:38	U	SMPL
J4806-03 - 1	RED-BLDG-3FL-BATH-SINK	2	0.0124	-245.1	PPB	07 Sep 2018 11:13:57	U	SMPL
J4807-01 - 1	15220032-1	2	0.0124	-245.1	PPB	07 Sep 2018 11:16:13	U	SMPL
PB112655BL - 1	PBW	-90	-0.0142	-245.1	PPB	07 Sep 2018 11:18:30	U	SMPL
PB112655BS - 1	LCSW	13112	3.8092	-245.1	PPB	07 Sep 2018 11:20:47	U	SMPL
J4798-01 - 1	090418-JETT-E-D-B	47	0.0255	-245.1	PPB	07 Sep 2018 11:23:05	U	SMPL
J4798-02 - 1	090418-JETT-E-D-S	-28	0.0037	-245.1	PPB	07 Sep 2018 11:25:24	U	SMPL
J4798-03 - 1	090418-JETT-E-U-B	-19	0.0063	-245.1	PPB	07 Sep 2018 11:27:42	U	SMPL
J4798-04 - 1	090418-JETT-E-U-S	36	0.0223	-245.1	PPB	07 Sep 2018 11:29:58	U	SMPL
CCV03 - 1	CCV03	17456	5.0673	-245.1	PPB	07 Sep 2018 11:32:14	U	SMPL
CCB03 - 1	CCB03	-145	-0.0301	-245.1	PPB	07 Sep 2018 11:34:30	U	SMPL
J4798-05 - 1	090418-SIDI-E-D-B	-51	-0.0029	-245.1	PPB	07 Sep 2018 11:36:48	U	SMPL
J4798-06 - 1	090418-SIDI-E-D-M	-42	-0.0003	-245.1	PPB	07 Sep 2018 11:39:04	U	SMPL
J4798-07 - 1	090418-SIDI-E-D-S	-9	0.0092	-245.1	PPB	07 Sep 2018 11:41:21	U	SMPL
J4798-08 - 1	090418-SIDI-E-U-B	-8	0.0095	-245.1	PPB	07 Sep 2018 11:43:37	U	SMPL
J4798-09 - 1	090418-SIDI-E-U-M	-43	-0.0006	-245.1	PPB	07 Sep 2018 11:45:54	U	SMPL
J4798-10 - 1	090418-SIDI-E-U-S	-7	0.0098	-245.1	PPB	07 Sep 2018 11:48:11	U	SMPL
J4798-11 - 1	090418-DBRI-E-D-B	-13	0.0081	-245.1	PPB	07 Sep 2018 11:50:29	U	SMPL
J4798-12 - 1	090418-DBRI-E-D-M	20	0.0176	-245.1	PPB	07 Sep 2018 11:52:46	U	SMPL
J4798-13 - 1	090418-DBRI-E-D-S	-17	0.0069	-245.1	PPB	07 Sep 2018 11:55:04	U	SMPL
J4798-14 - 1	090418-DBRI-E-U-B	29	0.0202	-245.1	PPB	07 Sep 2018 11:57:20	U	SMPL
CCV04 - 1	CCV04	17077	4.9575	-245.1	PPB	07 Sep 2018 11:59:36	U	SMPL
CCB04 - 1	CCB04	-223	-0.0527	-245.1	PPB	07 Sep 2018 12:01:53	U	SMPL
J4798-15 - 1	090418-DBRI-E-U-M	-25	0.0046	-245.1	PPB	07 Sep 2018 12:04:11	U	SMPL
J4798-16 - 1	090418-DBRI-E-U-S	0	0.0118	-245.1	PPB	07 Sep 2018 12:06:28	U	SMPL
J4798-17 - 1	090418-TIPC-E-D-M	-38	0.0008	-245.1	PPB	07 Sep 2018 12:08:44	U	SMPL
J4798-18 - 1	090418-TIPC-E-U-M	20	0.0176	-245.1	PPB	07 Sep 2018 12:11:01	U	SMPL
J4798-18DUP - 1	090418-TIPC-E-U-MDUP	96	0.0397	-245.1	PPB	07 Sep 2018 12:13:17	U	SMPL
J4798-18MS - 1	090418-TIPC-E-U-MMS	15492	4.4985	-245.1	PPB	07 Sep 2018 12:15:34	U	SMPL
J4798-18MSD - 1	090418-TIPC-E-U-MMSD	15046	4.3693	-245.1	PPB	07 Sep 2018 12:17:51	U	SMPL
CCV05 - 1	CCV05	17232	5.0024	-245.1	PPB	07 Sep 2018 12:20:09	U	SMPL
CCB05 - 1	CCB05	-178	-0.0397	-245.1	PPB	07 Sep 2018 12:22:29	U	SMPL
J4806-01LX5 - 1		-69	-0.0081	-245.1	PPB	07 Sep 2018 12:24:46	U	SMPL
J4806-01A - 1		15313	4.4466	-245.1	PPB	07 Sep 2018 12:27:03	U	SMPL
J4798-18LX5 - 1		-162	-0.0351	-245.1	PPB	07 Sep 2018 12:29:20	U	SMPL
J4798-18A - 1		17517	5.0849	-245.1	PPB	07 Sep 2018 12:31:38	U	SMPL
CCV06 - 1	CCV06	16975	4.9280	-245.1	PPB	07 Sep 2018 12:33:54	U	SMPL
CCB06 - 1	CCB06	-183	-0.0412	-245.1	PPB	07 Sep 2018 12:36:11	U	SMPL