


7471B

LB 97790
INSTRUMENT ID: CV1

Linear

 μ Abs.:

42600



A= 0.0000e+000

B= 2.3556e-004

C= -3.7440e-002

Rho= 0.9999923

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.025	-0.025	51	0.000	51				
0.2	0.200	0.210	0.010	1049	0.0 %	1049				
2.5	2.500	2.518	0.018	10850	0.0 %	10850				
5.0	5.000	5.009	0.009	21424	0.0 %	21424				
7.5	7.500	7.490	-0.010	31957	0.0 %	31957				
10.0	10.000	9.998	-0.002	42600	0.0 %	42600				

LB97790 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 07 Sep 2018 13:59:15

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	51	-	0.0000	7471B	PPB	07 Sep 2018 14:04:42	S	Std
0.2 - 1	50.2	1049	-	0.2000	7471B	PPB	07 Sep 2018 14:09:36	S	Std
2.5 - 1	52.5	10850	-	2.5000	7471B	PPB	07 Sep 2018 14:11:54	S	Std
5.0 - 1	55	21424	-	5.0000	7471B	PPB	07 Sep 2018 14:14:11	S	Std
7.5 - 1	57.5	31957	-	7.5000	7471B	PPB	07 Sep 2018 14:16:29	S	Std
10.0 - 1	510	42600	-	10.0000	7471B	PPB	07 Sep 2018 14:18:46	S	Std
ICV52 - 1	ICV52	16352	3.8145	-	7471B	PPB	07 Sep 2018 14:21:35	U	SMPL
ICB52 - 1	ICB52	135	-0.0056	-	7471B	PPB	07 Sep 2018 14:23:51	U	SMPL
CCV07 - 1	CCV07	21421	5.0086	-	7471B	PPB	07 Sep 2018 14:26:07	U	SMPL
CCB07 - 1	CCB07	120	-0.0092	-	7471B	PPB	07 Sep 2018 14:28:23	U	SMPL
CRA - 1	CRA	968	0.1906	-	7471B	PPB	07 Sep 2018 14:30:40	U	SMPL
HighStd - 1	HighStd	42832	10.0522	-	7471B	PPB	07 Sep 2018 14:32:55	U	SMPL
ChkStd - 1	ChkStd	30446	7.1345	-	7471B	PPB	07 Sep 2018 14:35:11	U	SMPL
PB112673BL - 1	PBS	110	-0.0115	-	7471B	PPB	07 Sep 2018 14:37:30	U	SMPL
PB112673BS - 1	LCSS	16373	3.8194	-	7471B	PPB	07 Sep 2018 14:39:48	U	SMPL
J4730-01 - 1	FK-BPM-02-008-ABCD	3177	0.7109	-	7471B	PPB	07 Sep 2018 14:42:05	U	SMPL
J4730-01DUP - 1	FK-BPM-02-008-ABCD DUP	3206	0.7178	-	7471B	PPB	07 Sep 2018 14:44:22	U	SMPL
J4730-01MS - 1	FK-BPM-02-008-ABCDMS	19250	4.4972	-	7471B	PPB	07 Sep 2018 14:46:40	U	SMPL
J4730-01MSD - 1	FK-BPM-02-008-ABCDMSD	19057	4.4517	-	7471B	PPB	07 Sep 2018 14:48:58	U	SMPL
J4774-01 - 1	OK-03-090618	2167	0.4730	-	7471B	PPB	07 Sep 2018 14:51:15	U	SMPL
CCV08 - 1	CCV08	21142	4.9429	-	7471B	PPB	07 Sep 2018 14:53:31	U	SMPL
CCB08 - 1	CCB08	183	0.0057	-	7471B	PPB	07 Sep 2018 14:55:47	U	SMPL
J4775-01 - 1	OR-02-090618-A	6205	1.4242	-	7471B	PPB	07 Sep 2018 14:58:03	U	SMPL
J4775-03 - 1	OR-02-090618-B	1897	0.4094	-	7471B	PPB	07 Sep 2018 15:00:19	U	SMPL
J4779-01 - 1	SU-03-090618-A	1396	0.2914	-	7471B	PPB	07 Sep 2018 15:02:35	U	SMPL
J4779-03 - 1	SU-03-090618-B	2248	0.4921	-	7471B	PPB	07 Sep 2018 15:04:51	U	SMPL
J4809-01 - 1	MS-DRUM-1-STONE	264	0.0247	-	7471B	PPB	07 Sep 2018 15:07:08	U	SMPL
J4809-05 - 1	MS-SLUDGE-V24129	104689	24.6235	-	7471B	PPB	07 Sep 2018 15:09:25	U	SMPL
J4825-17 - 1	EP1-WS-COMP-17	4774	1.0871	-	7471B	PPB	07 Sep 2018 15:11:42	U	SMPL
J4825-18 - 1	EP1-WS-COMP-18	5929	1.3592	-	7471B	PPB	07 Sep 2018 15:14:02	U	SMPL
J4730-01LX5 - 1		827	0.1574	-	7471B	PPB	07 Sep 2018 15:16:20	U	SMPL
J4730-01A - 1		21865	5.1132	-	7471B	PPB	07 Sep 2018 15:18:36	U	SMPL
CCV09 - 1	CCV09	21406	5.0050	-	7471B	PPB	07 Sep 2018 15:20:52	U	SMPL
CCB09 - 1	CCB09	216	0.0134	-	7471B	PPB	07 Sep 2018 15:23:08	U	SMPL
J4809-05DLX5 - 1		20724	4.8444	-	7471B	PPB	07 Sep 2018 15:25:25	U	SMPL
CCV10 - 1	CCV10	21447	5.0147	-	7471B	PPB	07 Sep 2018 15:27:41	U	SMPL
CCB10 - 1	CCB10	-133	-0.0688	-	7471B	PPB	07 Sep 2018 15:29:59	U	SMPL