



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB97793

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.1	4.01	09/07/2018	13:30
2	CAL2	1	Water			21.9	7.01	09/07/2018	13:32
3	CAL3	1	Water			21.8	10.04	09/07/2018	13:34
4	ICV	1	Water			21.9	7.00	09/07/2018	13:36
5	CCV1	1	Water			22.0	2.00	09/07/2018	13:38
6	J4828-01	1	Water			23.0	4.82	09/07/2018	13:40
7	J4828-01DUP	1	Water			23.1	4.83	09/07/2018	13:42
8	J4828-02	1	Water			23.1	4.28	09/07/2018	13:44
9	J4831-01	1	Water			14.9	5.72	09/07/2018	13:46
10	CCV2	1	Water			21.9	12.00	09/07/2018	13:48

WORKLIST(Hardcopy Internal Chain)

LB 97793

WorkList Name : pH090718

WorkList ID : 116506

Date : 9/7/2018 1:01:55 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/05/2018	Water	J4828-01 A	pH	Cool 4 deg C	STER14	I23	CHEMICAL-ANALYSIS-JS100	09/05/2018	SM 4500-PH B
09/05/2018	Water	J4828-02 A	pH	Cool 4 deg C	STER14	I23	CHEMICAL-ANALYSIS-MINI-C	09/05/2018	SM 4500-PH B
09/07/2018	Water	J4831-01 A	pH	Cool 4 deg C	MAJO01	I11	DRAIN-WATER-TANK-1	09/07/2018	SM 4500-PH B

Date/Time 09/07/18 1:10

Received by: SP

Relinquished by: CP

Date/Time 09/07/18 1:55

Received by: CP

Relinquished by: SP