



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB97796

Slope : 98.6

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.9	4.01	09/07/2018	15:40
2	CAL2	1	Water			21.8	7.01	09/07/2018	15:42
3	CAL3	1	Water			21.9	10.04	09/07/2018	15:44
4	ICV	1	Water			21.7	7.00	09/07/2018	15:46
5	CCV1	1	Water			22.0	2.00	09/07/2018	15:48
6	J4827-02	1	Water			19.2	7.08	09/07/2018	15:50
7	J4827-02DUP	1	Water			19.3	7.09	09/07/2018	15:52
8	CCV2	1	Water			21.9	12.00	09/07/2018	15:54

LB 97796

WORKLIST(Hardcopy Internal Chain)

WorkList Name : pH090718-1

WorkList ID : 116510

Date : 9/7/2018 3:33:55 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/06/2018	Water	J4827-02	pH	Cool 4 deg C	ENV/27	I23	DRUMS-WC	09/06/2018	9040C

Date/Time 09/07/18 3:35
Received by: SP
Relinquished by: CP

Date/Time 09/07/18 4:00
Received by: CP
Relinquished by: SP