

Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB97798

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.9	4.01	09/07/2018	15:40
2	CAL2	1	Water			21.8	7.01	09/07/2018	15:42
3	CAL3	1	Water			21.9	10.04	09/07/2018	15:44
4	ICV	1	Water			21.7	7.00	09/07/2018	15:46
5	CCV1	1	Water			21.8	2.00	09/07/2018	16:26
6	J4769-02	1	Solid	20.02	20	23.6	8.85	09/07/2018	16:28
7	J4769-02DUP	1	Solid	20.03	20	23.7	8.86	09/07/2018	16:30
8	J4769-04	1	Solid	20.02	20	23.9	9.03	09/07/2018	16:32
9	J4769-06	1	Solid	20.02	20	23.3	9.57	09/07/2018	16:34
10	J4776-02	1	Solid	20.04	20	23.8	8.49	09/07/2018	16:36
11	J4776-04	1	Solid	20.02	20	23.7	8.54	09/07/2018	16:38
12	J4777-02	1	Solid	20.02	20	23.7	8.74	09/07/2018	16:40
13	J4777-04	1	Solid	20.03	20	23.6	8.73	09/07/2018	16:42
14	J4777-06	1	Solid	20.04	20	23.7	8.23	09/07/2018	16:44
15	J4777-08	1	Solid	20.02	20	23.5	10.03	09/07/2018	16:46
16	CCV2	1	Water			21.9	12.00	09/07/2018	16:48
17	J4777-10	1	Solid	20.02	20	23.0	9.83	09/07/2018	16:50
18	J4821-07	1	Solid	20.03	20	23.2	6.95	09/07/2018	16:52
19	J4821-07DUP	1	Solid	20.04	20	23.3	6.96	09/07/2018	16:54
20	J4829-04	1	Solid	20.02	20	23.3	6.22	09/07/2018	16:56

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	J4829-08	1	Solid	20.03	20	28.8	7.31	09/07/2018	16:58
22	CCV3	1	Water			21.9	2.00	09/07/2018	17:00

WORKLIST(Hardcopy Internal Chain)

LB 97798

WorkList Name : corrosivity090718

WorkList ID : 116508

Date : 9/7/2018 3:14:44 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/07/2018	Solid	J4769-02	Corrosivity	Cool 4 deg C	PSEG05		HR-05-090718-A	09/07/2018	9045D
09/07/2018	Solid	J4769-04	Corrosivity	Cool 4 deg C	PSEG05		HR-05-090718-B	09/07/2018	9045D
09/07/2018	Solid	J4769-06	Corrosivity	Cool 4 deg C	PSEG05		HR-05-090718-C	09/07/2018	9045D
09/07/2018	Solid	J4776-02	Corrosivity	Cool 4 deg C	PSEG05		EO-02-090818-A	09/07/2018	9045D
09/07/2018	Solid	J4776-04	Corrosivity	Cool 4 deg C	PSEG05		EO-02-090818-B	09/07/2018	9045D
09/07/2018	Solid	J4777-02	Corrosivity	Cool 4 deg C	PSEG05	I21	PL-05-090718-A	09/07/2018	9045D
09/07/2018	Solid	J4777-04	Corrosivity	Cool 4 deg C	PSEG05	I21	PL-05-090718-B	09/07/2018	9045D
09/07/2018	Solid	J4777-06	Corrosivity	Cool 4 deg C	PSEG05	I21	PL-05-090718-C	09/07/2018	9045D
09/07/2018	Solid	J4777-08	Corrosivity	Cool 4 deg C	PSEG05	I21	PL-05-090718-D	09/07/2018	9045D
09/07/2018	Solid	J4777-10	Corrosivity	Cool 4 deg C	PSEG05	I21	PL-05-090718-E	09/07/2018	9045D
09/06/2018	Solid	J4821-07	Corrosivity	Cool 4 deg C	PARS15	I31	ENY-WC-S-01-0-62	09/06/2018	9045D
09/07/2018	Solid	J4829-04	Corrosivity	Cool 4 deg C	PSEG01	I31	OILY-STONES-PADS	09/07/2018	9045D
09/07/2018	Solid	J4829-08	Corrosivity	Cool 4 deg C	PSEG01	I31	OILY-STONES-3-4FT	09/07/2018	9045D

Date/Time 09/07/18 4:20

Received by: SP

Relinquished by: SP

Date/Time 09/07/18 5:05

Received by: SP

Relinquished by: SP