



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB97806
Analysis Date: 09/10/2018
BalanceID: WC SC-3
OvenID: EX OVEN-2

ANALYST: JIGNESH
REVIEWED BY: apatel
Extraction Date: 09/10/2018
Extraction IN Time: 08:10
Extraction OUT Time: 08:40

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB97806BL	LB97806BL	WATER	2	1000	100	3.1566	3.1566	3.02	3.1566	3.1566	0.0000	0
2	LB97806BS	LB97806BS	WATER	2	1000	100	3.0217	3.0217	3.01	3.0401	3.0401	0.0184	18.4
3	J4831-01	DRAIN-WATER-TANK-1	WATER	2	1000	100	3.0683	3.0683	3.02	3.0761	3.0761	0.0078	7.8
4	J4831-01MS	DRAIN-WATER-TANK-1	WATER	2	1000	100	3.0807	3.0807	3.04	3.1109	3.1109	0.0302	30.2
5	J4831-01MSD	DRAIN-WATER-TANK-1	WATER	2	1000	100	3.0161	3.0161	3.05	3.0465	3.0465	0.0304	30.4



QC Batch# LB97806 Test: TPH Analysis Date: 09/10/2018

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2408
pH Paper 0-14	W2373
Sodium Sulfate	EP1876
1:1 HCL	WP64228
Silica Gel	W2199
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP64229
LCSWD	NA	NA
MS/MSD	10.00 ML	WP64230

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance:	<u>0.0018</u>	In OVEN TEMP :	<u>70</u>	Dessicator Time In :	<u>10:16</u>
1.0000 gram Balance:	<u>1.0005</u>	In Time:	<u>09:15</u>		
Bal Check Time:	<u>08:30</u>	Out OVEN TEMP:	<u>70</u>	Dessicator Time Out:	<u>10:40</u>
		Out Time:	<u>10:15</u>		

After Analysis

0.0020 gram Balance:	<u>0.0021</u>	In OVEN TEMP :	<u>71</u>	Dessicator Time In :	<u>11:41</u>
1.0000 gram Balance:	<u>1.0003</u>	In Time:	<u>11:15</u>		
Bal Check Time:	<u>12:45</u>	Out OVEN TEMP:	<u>71</u>	Dessicator Time Out:	<u>12:15</u>
		Out Time:	<u>11:40</u>		

LB94806

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TPH J4831

WorkList ID : 116519

Date : 9/10/2018 8:04:27 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/14/2018	Water	J4831-01	TPH	Conc H2SO4 to	MAJO01	I11	DRAIN-WATER-TANK-1	09/07/2018	1664A

Ca, F, H,

Date/Time 09-10-18 8:05 AM
Received by: JB
Relinquished by: CB

Date/Time 09-10-18 10:30 AM
Received by: CB
Relinquished by: JB