


245.1

LB 97811
INSTRUMENT ID: CV1

Linear

 μ Abs.:

32787



A= 0.0000e+000

B= 3.0230e-004

C= 1.1651e-002

Rho= 0.9998732

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.009	0.009	-8	0.000	-8				
0.05	0.050	0.058	0.008	154	0.0 %	154				
0.2	0.200	0.204	0.004	636	0.0 %	636				
2.5	2.500	2.424	-0.076	7980	0.0 %	7980				
5.0	5.000	5.021	0.021	16572	0.0 %	16572				
7.5	7.500	7.610	0.110	25136	0.0 %	25136				
10.0	10.000	9.923	-0.077	32787	0.0 %	32787				

LB97811 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 10 Sep 2018 11:15:30

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-8	-	0.0000	245.1	PPB	10 Sep 2018 11:16:53	S	Std
0.05 - 1	50.5	154	-	0.0500	245.1	PPB	10 Sep 2018 11:19:10	S	Std
0.2 - 1	50.2	636	-	0.2000	245.1	PPB	10 Sep 2018 11:21:28	S	Std
2.5 - 1	52.5	7980	-	2.5000	245.1	PPB	10 Sep 2018 11:23:45	S	Std
5.0 - 1	55	16572	-	5.0000	245.1	PPB	10 Sep 2018 11:26:03	S	Std
7.5 - 1	57.5	25136	-	7.5000	245.1	PPB	10 Sep 2018 11:30:55	S	Std
10.0 - 1	510	32787	-	10.0000	245.1	PPB	10 Sep 2018 11:38:33	S	Std
ICV54 - 1	ICV54	13050	3.9566	-245.1		PPB	10 Sep 2018 11:41:17	U	SMPL
ICB54 - 1	ICB54	-280	-0.0730	-245.1		PPB	10 Sep 2018 11:43:34	U	SMPL
CCV14 - 1	CCV14	16749	5.0748	-245.1		PPB	10 Sep 2018 11:45:52	U	SMPL
CCB14 - 1	CCB14	-196	-0.0476	-245.1		PPB	10 Sep 2018 11:48:07	U	SMPL
CRA - 1	CRA	101	0.0422	-245.1		PPB	10 Sep 2018 11:50:25	U	SMPL
HighStd - 1	HighStd	32204	9.7469	-245.1		PPB	10 Sep 2018 11:52:41	U	SMPL
ChkStd - 1	ChkStd	23703	7.1770	-245.1		PPB	10 Sep 2018 11:54:57	U	SMPL
PB112699BL - 1	PBW	-147	-0.0328	-245.1		PPB	10 Sep 2018 11:57:16	U	SMPL
PB112699BS - 1	LCSW	13060	3.9597	-245.1		PPB	10 Sep 2018 12:02:15	U	SMPL
J4811-01 - 1	090518-DBRI-E-D-B	-74	-0.0107	-245.1		PPB	10 Sep 2018 12:04:32	U	SMPL
J4811-02 - 1	090518-DBRI-E-D-M	248	0.0866	-245.1		PPB	10 Sep 2018 12:06:51	U	SMPL
J4811-03 - 1	090518-DBRI-E-D-S	249	0.0869	-245.1		PPB	10 Sep 2018 12:11:54	U	SMPL
J4811-04 - 1	090518-DBRI-E-U-B	268	0.0927	-245.1		PPB	10 Sep 2018 12:14:13	U	SMPL
J4811-05 - 1	090518-DBRI-E-U-M	248	0.0866	-245.1		PPB	10 Sep 2018 12:16:28	U	SMPL
CCV15 - 1	CCV15	16486	4.9953	-245.1		PPB	10 Sep 2018 12:18:44	U	SMPL
CCB15 - 1	CCB15	99	0.0416	-245.1		PPB	10 Sep 2018 12:23:44	U	SMPL
J4811-06 - 1	090518-DBRI-E-U-S	225	0.0797	-245.1		PPB	10 Sep 2018 12:26:00	U	SMPL
J4811-07 - 1	090518-SIDI-E-D-B	263	0.0912	-245.1		PPB	10 Sep 2018 12:28:16	U	SMPL
J4811-08 - 1	090518-SIDI-E-D-S	354	0.1187	-245.1		PPB	10 Sep 2018 12:30:32	U	SMPL
J4811-09 - 1	090518-SIDI-E-U-B	318	0.1078	-245.1		PPB	10 Sep 2018 12:32:49	U	SMPL
J4811-10 - 1	090518-SIDI-E-U-S	195	0.0706	-245.1		PPB	10 Sep 2018 12:35:05	U	SMPL
J4811-11 - 1	090518-TIPC-E-D-M	218	0.0776	-245.1		PPB	10 Sep 2018 12:37:22	U	SMPL
J4811-12 - 1	090518-TIPC-E-U-M	225	0.0797	-245.1		PPB	10 Sep 2018 12:39:39	U	SMPL
J4814-01 - 1	32220008-3	179	0.0658	-245.1		PPB	10 Sep 2018 12:41:57	U	SMPL
J4814-01DUP - 1	32220008-3DUP	219	0.0779	-245.1		PPB	10 Sep 2018 12:46:50	U	SMPL
J4814-01MS - 1	32220008-3MS	13058	3.9591	-245.1		PPB	10 Sep 2018 12:49:06	U	SMPL
CCV16 - 1	CCV16	16660	5.0479	-245.1		PPB	10 Sep 2018 12:51:22	U	SMPL
CCB16 - 1	CCB16	-44	-0.0017	-245.1		PPB	10 Sep 2018 12:56:24	U	SMPL
J4814-01MSD - 1	32220008-3MSD	13204	4.0032	-245.1		PPB	10 Sep 2018 12:58:39	U	SMPL
J4814-01LX5 - 1		37	0.0228	-245.1		PPB	10 Sep 2018 13:00:56	U	SMPL
J4814-01A - 1		13262	4.0207	-245.1		PPB	10 Sep 2018 13:03:13	U	SMPL
CCV17 - 1	CCV17	15737	4.7689	-245.1		PPB	10 Sep 2018 13:05:30	U	SMPL
CCB17 - 1	CCB17	-184	-0.0440	-245.1		PPB	10 Sep 2018 13:07:48	U	SMPL