



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB97815

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.0	80.6	80.6	09/10/2018	10:25
2	J4827-02		1	9	>100.0	>212.0	>212.0	09/10/2018	10:55
3	J4827-02DUP		1	9	>100.0	>212.0	>212.0	09/10/2018	11:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB97815

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp91018 WorkList ID : 116545 Date : 9/10/2018 9:47:26 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/20/2018	Water	J4827-02 B	Flash Point	Cool 4 deg C	ENV127	I23	DRUMS-WC	09/06/2018	1010A

Date/Time 09/10/18 9:50 AM
Received by: KS
Relinquished by: CP

Date/Time 09/10/18 11:30 AM
Received by: CP
Relinquished by: KS