


7471B

LB 97819
INSTRUMENT ID: CV2

Linear

 μ Abs.:

44134



A= 0.0000e+000

B= 2.2857e-004

C= -7.7842e-002

Rho= 0.9998975

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.052	-0.052	115	0.000	115				
0.2	0.200	0.258	0.058	1468	0.0 %	1468				
2.5	2.500	2.546	0.046	11478	0.0 %	11478				
5.0	5.000	4.911	-0.089	21828	0.0 %	21828				
7.5	7.500	7.527	0.027	33271	0.0 %	33271				
10.0	10.000	10.010	0.010	44134	0.0 %	44134				

LB97819 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 10 Sep 2018 15:31:12

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	115	-	0.00007471B	PPB		10 Sep 2018 15:33:57	S	Std
0.2 - 1	50.2	1468	-	0.20007471B	PPB		10 Sep 2018 15:36:15	S	Std
2.5 - 1	52.5	11478	-	2.50007471B	PPB		10 Sep 2018 15:38:33	S	Std
5.0 - 1	55	21828	-	5.00007471B	PPB		10 Sep 2018 15:40:50	S	Std
7.5 - 1	57.5	33271	-	7.50007471B	PPB		10 Sep 2018 15:43:07	S	Std
10.0 - 1	58	44134	-	10.00007471B	PPB		10 Sep 2018 15:45:25	S	Std
ICV55 - 1	ICV55	16902	3.7855	-7471B	PPB		10 Sep 2018 15:48:12	U	SMPL
ICB55 - 1	ICB55	24	-0.0724	-7471B	PPB		10 Sep 2018 15:50:28	U	SMPL
CCV18 - 1	CCV18	22076	4.9681	-7471B	PPB		10 Sep 2018 15:52:44	U	SMPL
CCB18 - 1	CCB18	13	-0.0749	-7471B	PPB		10 Sep 2018 15:55:00	U	SMPL
CRA - 1	CRA	1400	0.2422	-7471B	PPB		10 Sep 2018 15:57:15	U	SMPL
HighStd - 1	HighStd	44231	10.0320	-7471B	PPB		10 Sep 2018 15:59:31	U	SMPL
ChkStd - 1	ChkStd	31104	7.0316	-7471B	PPB		10 Sep 2018 16:01:47	U	SMPL
PB112706BL - 1	PBS	-9	-0.0799	-7471B	PPB		10 Sep 2018 16:09:33	U	SMPL
PB112706BS - 1	LCSS	16682	3.7352	-7471B	PPB		10 Sep 2018 16:11:49	U	SMPL
J4769-01 - 1	HR-05-090718-A	7920	1.7324	-7471B	PPB		10 Sep 2018 16:14:06	U	SMPL
J4769-01DUP - 1	HR-05-090718-ADUP	8045	1.7610	-7471B	PPB		10 Sep 2018 16:18:59	U	SMPL
J4769-01MS - 1	HR-05-090718-AMS	25225	5.6878	-7471B	PPB		10 Sep 2018 16:21:16	U	SMPL
J4769-01MSD - 1	HR-05-090718-AMSD	26094	5.8865	-7471B	PPB		10 Sep 2018 16:23:34	U	SMPL
J4769-03 - 1	HR-05-090718-B	5404	1.1574	-7471B	PPB		10 Sep 2018 16:25:50	U	SMPL
CCV19 - 1	CCV19	21813	4.9080	-7471B	PPB		10 Sep 2018 16:28:07	U	SMPL
CCB19 - 1	CCB19	-47	-0.0886	-7471B	PPB		10 Sep 2018 16:30:23	U	SMPL
J4769-05 - 1	HR-05-090718-C	7332	1.5980	-7471B	PPB		10 Sep 2018 16:32:39	U	SMPL
J4776-01 - 1	EO-02-090818-A	3815	0.7942	-7471B	PPB		10 Sep 2018 16:34:55	U	SMPL
J4776-03 - 1	EO-02-090818-B	7620	1.6639	-7471B	PPB		10 Sep 2018 16:37:11	U	SMPL
J4777-01 - 1	PL-05-090718-A	1021	0.1555	-7471B	PPB		10 Sep 2018 16:39:27	U	SMPL
J4777-03 - 1	PL-05-090718-B	1511	0.2675	-7471B	PPB		10 Sep 2018 16:41:44	U	SMPL
J4777-05 - 1	PL-05-090718-C	1332	0.2266	-7471B	PPB		10 Sep 2018 16:44:01	U	SMPL
J4777-07 - 1	PL-05-090718-D	4100	0.8593	-7471B	PPB		10 Sep 2018 16:46:18	U	SMPL
J4777-09 - 1	PL-05-090718-E	2259	0.4385	-7471B	PPB		10 Sep 2018 16:48:36	U	SMPL
J4821-05 - 1	ENY-SB-01-02-03-0-62	303	-0.0086	-7471B	PPB		10 Sep 2018 16:50:54	U	SMPL
J4821-06 - 1	ENY-BC-02	38208	8.6554	-7471B	PPB		10 Sep 2018 16:53:10	U	SMPL
CCV20 - 1	CCV20	21661	4.8732	-7471B	PPB		10 Sep 2018 16:55:26	U	SMPL
CCB20 - 1	CCB20	20	-0.0733	-7471B	PPB		10 Sep 2018 16:57:44	U	SMPL
J4829-01 - 1	OILY-STONES-PADS	6151	1.3281	-7471B	PPB		10 Sep 2018 16:59:59	U	SMPL
J4829-05 - 1	OILY-STONES-3-4FT	905	0.1290	-7471B	PPB		10 Sep 2018 17:02:15	U	SMPL
J4834-01 - 1	C-1	1497	0.2643	-7471B	PPB		10 Sep 2018 17:04:31	U	SMPL
J4836-01 - 1	FK-GF-02-025-A	2297	0.4472	-7471B	PPB		10 Sep 2018 17:06:47	U	SMPL
J4836-03 - 1	FK-GF-02-025-B	2639	0.5254	-7471B	PPB		10 Sep 2018 17:09:04	U	SMPL
J4769-01LX5 - 1		1589	0.2854	-7471B	PPB		10 Sep 2018 17:11:21	U	SMPL
J4769-01A - 1		25485	5.7473	-7471B	PPB		10 Sep 2018 17:13:38	U	SMPL
CCV21 - 1	CCV21	21632	4.8666	-7471B	PPB		10 Sep 2018 17:15:55	U	SMPL
CCB21 - 1	CCB21	38	-0.0692	-7471B	PPB		10 Sep 2018 17:18:14	U	SMPL