

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 09/11/2018

TEMP1 IN: 103 09/11/2018 07:00 TEMP OUT: 104 09/11/2018 08:05

TEMP2 IN: 104 09/11/2018 09:00 TEMP OUT: 104 09/11/2018 10:00

TEMP3 IN: 105 09/11/2018 11:30 TEMP OUT: 105 09/11/2018 13:10

TEMP4 IN: 105 09/11/2018 14:00 TEMP OUT: 105 09/11/2018 15:36

Run Number: LB97825

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB97825BL	LB97825BL	1.4155	1.4155	100	1.4155	1.4155	1.4155	0.0000	0
2	LB97825BS	LB97825BS	1.4087	1.4087	100	1.4593	1.4593	1.4593	0.0506	506
3	J4841-01	090718-DBRI-F-D-B	1.4094	1.4094	200	1.4245	1.4245	1.4245	0.0151	75.5
4	J4841-01DUP	090718-DBRI-F-D-BDUP	1.4159	1.4159	200	1.4312	1.4312	1.4312	0.0153	76.5
5	J4841-02	090718-DBRI-F-D-M	1.4006	1.4006	300	1.4127	1.4127	1.4127	0.0121	40.3
6	J4841-03	090718-DBRI-F-D-S	1.4136	1.4136	400	1.4222	1.4222	1.4222	0.0086	21.5
7	J4841-04	090718-DBRI-F-U-B	1.4042	1.4042	200	1.4308	1.4308	1.4308	0.0266	133
8	J4841-05	090718-DBRI-F-U-M	1.4121	1.4121	500	1.4273	1.4273	1.4273	0.0152	30.4
9	J4841-06	090718-DBRI-F-U-S	1.4022	1.4022	400	1.4102	1.4102	1.4102	0.0080	20
10	J4841-07	090718-SIDI-F-D-B	1.3919	1.3919	200	1.4146	1.4146	1.4146	0.0227	113.5
11	J4841-08	090718-SIDI-F-D-M	1.3951	1.3951	400	1.4056	1.4056	1.4056	0.0105	26.3
12	J4841-09	090718-SIDI-F-D-S	1.3995	1.3995	400	1.4088	1.4088	1.4088	0.0093	23.3
13	J4841-10	090718-SIDI-F-U-B	1.4118	1.4118	300	1.4333	1.4333	1.4333	0.0215	71.7
14	J4841-11	090718-SIDI-F-U-M	1.4033	1.4033	400	1.4174	1.4174	1.4174	0.0141	35.2
15	J4841-12	090718-SIDI-F-U-S	1.3938	1.3938	200	1.4010	1.4010	1.4010	0.0072	36

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 09/11/2018

TEMP1 IN: 103 09/11/2018 07:00 TEMP OUT: 104 09/11/2018 08:05

TEMP2 IN: 104 09/11/2018 09:00 TEMP OUT: 104 09/11/2018 10:00

TEMP3 IN: 105 09/11/2018 11:30 TEMP OUT: 105 09/11/2018 13:10

TEMP4 IN: 105 09/11/2018 14:00 TEMP OUT: 105 09/11/2018 15:36

Run Number: LB97825

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	J4841-12DUP	090718-SIDI-F-U-SDUP	1.4046	1.4046	200	1.4115	1.4115	1.4115	0.0069	34.5
17	J4844-01	001A	1.4027	1.4027	500	1.4198	1.4198	1.4198	0.0171	34.2
18	J4844-02	001D	1.3973	1.3973	900	1.3977	1.3977	1.3977	0.0004	0.4
19	J4844-03	006	1.4013	1.4013	300	1.4065	1.4065	1.4065	0.0052	17.3
20	J4844-05	RAINSY-AFTER	1.4019	1.4019	900	1.4037	1.4037	1.4037	0.0018	2

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS091118

WorkList ID : 116561

Date : 9/11/2018 7:40:28 AM

2197825

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/14/2018	Water	J4841-01 C	TSS	Cool 4 deg C	HDR103	I13	090718-DBRI-F-D-B	09/07/2018	SM2540 D
09/14/2018	Water	J4841-02 C	TSS	Cool 4 deg C	HDR103	I13	090718-DBRI-F-D-M	09/07/2018	SM2540 D
09/14/2018	Water	J4841-03 C	TSS	Cool 4 deg C	HDR103	I13	090718-DBRI-F-D-S	09/07/2018	SM2540 D
09/14/2018	Water	J4841-04 C	TSS	Cool 4 deg C	HDR103	I13	090718-DBRI-F-U-B	09/07/2018	SM2540 D
09/14/2018	Water	J4841-05 C	TSS	Cool 4 deg C	HDR103	I13	090718-DBRI-F-U-M	09/07/2018	SM2540 D
09/14/2018	Water	J4841-06 C	TSS	Cool 4 deg C	HDR103	I13	090718-DBRI-F-U-S	09/07/2018	SM2540 D
09/14/2018	Water	J4841-07 C	TSS	Cool 4 deg C	HDR103	I13	090718-SIDI-F-D-B	09/07/2018	SM2540 D
09/14/2018	Water	J4841-08 C	TSS	Cool 4 deg C	HDR103	I13	090718-SIDI-F-D-M	09/07/2018	SM2540 D
09/14/2018	Water	J4841-09 C	TSS	Cool 4 deg C	HDR103	I13	090718-SIDI-F-D-S	09/07/2018	SM2540 D
09/14/2018	Water	J4841-10 C	TSS	Cool 4 deg C	HDR103	I13	090718-SIDI-F-U-B	09/07/2018	SM2540 D
09/14/2018	Water	J4841-11 C	TSS	Cool 4 deg C	HDR103	I13	090718-SIDI-F-U-M	09/07/2018	SM2540 D
09/14/2018	Water	J4841-12 C	TSS	Cool 4 deg C	HDR103	I13	090718-SIDI-F-U-S	09/07/2018	SM2540 D
09/14/2018	Water	J4844-01 A	TSS	Cool 4 deg C	FOST03	I11	001A	09/07/2018	SM2540 D
09/14/2018	Water	J4844-02 A	TSS	Cool 4 deg C	FOST03	I11	001D	09/07/2018	SM2540 D
09/14/2018	Water	J4844-03 A	TSS	Cool 4 deg C	FOST03	I11	006	09/07/2018	SM2540 D
09/14/2018	Water	J4844-05 A	TSS	Cool 4 deg C	FOST03	I11	RAINSY-AFTER	09/07/2018	SM2540 D

FP-7042602

Date/Time 09/14/18 08:00
 Received by: AS
 Relinquished by: CQ

Date/Time 09/14/18 12:00
 Received by: 25
 Relinquished by: 25