


7470A

LA97827
INSTRUMENT ID: CV2

Linear

 μ Abs.:

37179



A= 0.0000e+000

B= 2.7268e-004

C= -2.7779e-002

Rho= 0.9996880

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.024	-0.024	15	0.000	15				
0.2	0.200	0.213	0.013	884	0.0 %	884				
2.5	2.500	2.593	0.093	9611	0.0 %	9611				
5.0	5.000	4.977	-0.023	18354	0.0 %	18354				
7.5	7.500	7.330	-0.170	26983	0.0 %	26983				
10.0	10.000	10.110	0.110	37179	0.0 %	37179				

LB97827 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 11 Sep 2018 08:32:05

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	15	-	0.00007470A	PPB		11 Sep 2018 08:36:14	S	Std
0.2 - 1	502	884	-	0.20007470A	PPB		11 Sep 2018 08:38:22	S	Std
2.5 - 1	52.5	9611	-	2.50007470A	PPB		11 Sep 2018 08:40:32	S	Std
5.0 - 1	55	18354	-	5.00007470A	PPB		11 Sep 2018 08:42:40	S	Std
7.5 - 1	57.5	26983	-	7.50007470A	PPB		11 Sep 2018 08:44:48	S	Std
10.0 - 1	510	37179	-	10.00007470A	PPB		11 Sep 2018 08:46:57	S	Std
ICV56 - 1	ICV56	14702	3.9812	-7470A	PPB		11 Sep 2018 08:49:34	U	SMPL
ICB56 - 1	ICB56	-9	-0.0302	-7470A	PPB		11 Sep 2018 08:51:40	U	SMPL
CCV22 - 1	CCV22	18746	5.0839	-7470A	PPB		11 Sep 2018 08:53:48	U	SMPL
CCB22 - 1	CCB22	7	-0.0259	-7470A	PPB		11 Sep 2018 08:55:55	U	SMPL
CRA - 1	CRA	989	0.2419	-7470A	PPB		11 Sep 2018 08:58:03	U	SMPL
HighStd - 1	HighStd	37379	10.1649	-7470A	PPB		11 Sep 2018 09:00:12	U	SMPL
ChkStd - 1	ChkStd	25743	6.9919	-7470A	PPB		11 Sep 2018 09:02:20	U	SMPL
PB112744BL - 1	PBW	31	-0.0193	-7470A	PPB		11 Sep 2018 09:07:01	U	SMPL
PB112744BS - 1	LCSW	14064	3.8072	-7470A	PPB		11 Sep 2018 09:09:08	U	SMPL
PB112744TB - 1	PB112703TB	-61	-0.0444	-7470A	PPB		11 Sep 2018 09:11:16	U	SMPL
J4769-02 - 1	HR-05-090718-A	116	0.0039	-7470A	PPB		11 Sep 2018 09:13:24	U	SMPL
J4769-04 - 1	HR-05-090718-B	97	-0.0013	-7470A	PPB		11 Sep 2018 09:15:32	U	SMPL
J4769-06 - 1	HR-05-090718-C	61	-0.0111	-7470A	PPB		11 Sep 2018 09:17:41	U	SMPL
J4776-02 - 1	EO-02-090818-A	159	0.0156	-7470A	PPB		11 Sep 2018 09:19:47	U	SMPL
CCV23 - 1	CCV23	18776	5.0921	-7470A	PPB		11 Sep 2018 09:21:54	U	SMPL
CCB23 - 1	CCB23	-59	-0.0439	-7470A	PPB		11 Sep 2018 09:24:01	U	SMPL
J4776-04 - 1	EO-02-090818-B	75	-0.0073	-7470A	PPB		11 Sep 2018 09:26:08	U	SMPL
J4777-02 - 1	PL-05-090718-A	107	0.0014	-7470A	PPB		11 Sep 2018 09:28:15	U	SMPL
J4777-04 - 1	PL-05-090718-B	148	0.0126	-7470A	PPB		11 Sep 2018 09:30:21	U	SMPL
J4777-06 - 1	PL-05-090718-C	169	0.0183	-7470A	PPB		11 Sep 2018 09:32:28	U	SMPL
J4777-08 - 1	PL-05-090718-D	128	0.0071	-7470A	PPB		11 Sep 2018 09:34:36	U	SMPL
J4777-10 - 1	PL-05-090718-E	190	0.0240	-7470A	PPB		11 Sep 2018 09:36:43	U	SMPL
J4821-07 - 1	ENY-WC-S-01-0-62	104	0.0006	-7470A	PPB		11 Sep 2018 09:38:51	U	SMPL
J4829-04 - 1	OILY-STONES-PADS	151	0.0134	-7470A	PPB		11 Sep 2018 09:41:00	U	SMPL
J4829-08 - 1	OILY-STONES-3-4FT	47	-0.0150	-7470A	PPB		11 Sep 2018 09:43:08	U	SMPL
J4829-08DUP - 1	OILY-STONES-3-4FTDUP	98	-0.0011	-7470A	PPB		11 Sep 2018 09:45:15	U	SMPL
CCV24 - 1	CCV24	18441	5.0008	-7470A	PPB		11 Sep 2018 09:47:22	U	SMPL
CCB24 - 1	CCB24	62	-0.0109	-7470A	PPB		11 Sep 2018 09:49:29	U	SMPL
J4829-08MS - 1	OILY-STONES-3-4FTMS	13997	3.7890	-7470A	PPB		11 Sep 2018 09:51:35	U	SMPL
J4829-08MSD - 1	OILY-STONES-3-4FTMSD	13826	3.7423	-7470A	PPB		11 Sep 2018 09:53:42	U	SMPL
PB112745BL - 1	PBW	8	-0.0256	-7470A	PPB		11 Sep 2018 09:55:49	U	SMPL
PB112745BS - 1	LCSW	13693	3.7061	-7470A	PPB		11 Sep 2018 09:57:57	U	SMPL
PB112745TB - 1	PB112705TB	52	-0.0136	-7470A	PPB		11 Sep 2018 10:00:04	U	SMPL
J4827-02 - 1	DRUMS-WC	179	0.0210	-7470A	PPB		11 Sep 2018 10:02:14	U	SMPL
J4827-02DUP - 1	DRUMS-WCDUP	206	0.0284	-7470A	PPB		11 Sep 2018 10:04:22	U	SMPL
J4827-02MS - 1	DRUMS-WCMS	13804	3.7363	-7470A	PPB		11 Sep 2018 10:06:30	U	SMPL
J4827-02MSD - 1	DRUMS-WCMSD	13138	3.5547	-7470A	PPB		11 Sep 2018 10:08:39	U	SMPL
CCV25 - 1	CCV25	18764	5.0889	-7470A	PPB		11 Sep 2018 10:10:46	U	SMPL
CCB25 - 1	CCB25	-87	-0.0515	-7470A	PPB		11 Sep 2018 10:12:54	U	SMPL
J4829-08LX5 - 1		245	0.0390	-7470A	PPB		11 Sep 2018 10:15:01	U	SMPL
J4829-08A - 1		14900	4.0352	-7470A	PPB		11 Sep 2018 10:17:09	U	SMPL
J4827-02LX5 - 1		187	0.0232	-7470A	PPB		11 Sep 2018 10:19:16	U	SMPL
J4827-02A - 1		14015	3.7939	-7470A	PPB		11 Sep 2018 10:21:24	U	SMPL
CCV26 - 1	CCV26	18619	5.0493	-7470A	PPB		11 Sep 2018 10:23:31	U	SMPL
CCB26 - 1	CCB26	-86	-0.0512	-7470A	PPB		11 Sep 2018 10:25:39	U	SMPL