



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB97828

Slope : 99.1

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.8	4.01	09/10/2018	16:00
2	CAL2	1	Water			20.9	7.01	09/10/2018	16:02
3	CAL3	1	Water			21.1	10.04	09/10/2018	16:04
4	ICV	1	Water			21.0	7.00	09/10/2018	16:06
5	CCV1	1	Water			20.8	2.00	09/10/2018	16:08
6	J4843-01	1	Water			18.4	5.16	09/10/2018	16:10
7	J4843-01DUP	1	Water			18.5	5.17	09/10/2018	16:12
8	CCV2	1	Water			21.0	12.00	09/10/2018	16:14

WORKLIST(Hardcopy Internal Chain)

LB97828

WorkList Name : pH091018

WorkList ID : 116557

Date : 9/10/2018 3:49:08 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/07/2018	Water	J4843-01	pH	Cool 4 deg C	STER04	I22	DEIONIZED-POOL-WATER	09/07/2018	9040C

Date/Time 09/10/18 3:55
 Received by: SP
 Relinquished by: CP

Date/Time 09/10/18 4:20
 Received by: CP
 Relinquished by: SP