

245.1


LB97833

INSTRUMENT ID: CV1

Linear

 μ Abs.:

33774



A= 0.0000e+000

B= 2.9507e-004

C= -9.0971e-002

Rho= 0.9997926

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.008	-0.008	280	0.000	280				
0.05	0.050	0.037	-0.013	432	0.0 %	432				
0.2	0.200	0.165	-0.035	869	0.0 %	869				
2.5	2.500	2.525	0.025	8864	0.0 %	8864				
5.0	5.000	5.008	0.008	17280	0.0 %	17280				
7.5	7.500	7.649	0.149	26231	0.0 %	26231				
10.0	10.000	9.875	-0.125	33774	0.0 %	33774				

LB97833 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 11 Sep 2018 11:27:25

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	280	-	0.0000	245.1	PPB	11 Sep 2018 11:56:51	S	Std
0.05 - 1	50.05	432	-	0.0500	245.1	PPB	11 Sep 2018 11:59:09	S	Std
0.2 - 1	50.2	869	-	0.2000	245.1	PPB	11 Sep 2018 12:01:27	S	Std
2.5 - 1	52.5	8864	-	2.5000	245.1	PPB	11 Sep 2018 12:03:44	S	Std
5.0 - 1	55	17280	-	5.0000	245.1	PPB	11 Sep 2018 12:06:02	S	Std
7.5 - 1	57.5	26231	-	7.5000	245.1	PPB	11 Sep 2018 12:08:18	S	Std
10.0 - 1	510	33774	-	10.0000	245.1	PPB	11 Sep 2018 12:13:15	S	Std
ICV57 - 1	ICV57	13928	4.0188	-245.1		PPB	11 Sep 2018 12:16:02	U	SMPL
ICB57 - 1	ICB57	5	-0.0895	-245.1		PPB	11 Sep 2018 12:18:18	U	SMPL
CCV27 - 1	CCV27	17571	5.0938	-245.1		PPB	11 Sep 2018 12:20:37	U	SMPL
CCB27 - 1	CCB27	6	-0.0892	-245.1		PPB	11 Sep 2018 12:22:52	U	SMPL
CRA - 1	CRA	493	0.0545	-245.1		PPB	11 Sep 2018 12:25:09	U	SMPL
HighStd - 1	HighStd	34110	9.9740	-245.1		PPB	11 Sep 2018 12:27:25	U	SMPL
ChkStd - 1	ChkStd	25057	7.3027	-245.1		PPB	11 Sep 2018 12:29:41	U	SMPL
PB112746BL - 1	PBW	198	-0.0325	-245.1		PPB	11 Sep 2018 12:34:37	U	SMPL
PB112746BS - 1	LCSW	13660	3.9397	-245.1		PPB	11 Sep 2018 12:36:53	U	SMPL
J4830-01 - 1	090618-SIDI-F-D-B	200	-0.0320	-245.1		PPB	11 Sep 2018 12:39:10	U	SMPL
J4830-02 - 1	090618-SIDI-F-D-M	329	0.0061	-245.1		PPB	11 Sep 2018 12:41:28	U	SMPL
J4830-03 - 1	090618-SIDI-F-D-S	216	-0.0272	-245.1		PPB	11 Sep 2018 12:43:45	U	SMPL
J4830-04 - 1	090618-SIDI-F-U-B	253	-0.0163	-245.1		PPB	11 Sep 2018 12:46:04	U	SMPL
J4830-05 - 1	090618-SIDI-F-U-M	222	-0.0255	-245.1		PPB	11 Sep 2018 12:48:20	U	SMPL
CCV28 - 1	CCV28	17245	4.9976	-245.1		PPB	11 Sep 2018 12:50:35	U	SMPL
CCB28 - 1	CCB28	-4	-0.0922	-245.1		PPB	11 Sep 2018 12:52:52	U	SMPL
J4830-06 - 1	090618-SIDI-F-U-S	153	-0.0458	-245.1		PPB	11 Sep 2018 12:55:10	U	SMPL
J4830-07 - 1	090618-TIPC-E-D-M	219	-0.0264	-245.1		PPB	11 Sep 2018 12:57:25	U	SMPL
J4830-08 - 1	090618-TIPC-E-U-M	218	-0.0266	-245.1		PPB	11 Sep 2018 12:59:42	U	SMPL
J4830-09 - 1	090618-DBRI-E-D-B	247	-0.0181	-245.1		PPB	11 Sep 2018 13:01:58	U	SMPL
J4830-10 - 1	090618-DBRI-E-D-M	162	-0.0432	-245.1		PPB	11 Sep 2018 13:04:15	U	SMPL
J4830-11 - 1	090618-DBRI-E-D-S	104	-0.0603	-245.1		PPB	11 Sep 2018 13:06:32	U	SMPL
J4830-12 - 1	090618-DBRI-E-U-B	176	-0.0390	-245.1		PPB	11 Sep 2018 13:08:49	U	SMPL
J4830-13 - 1	090618-DBRI-E-U-M	188	-0.0355	-245.1		PPB	11 Sep 2018 13:11:07	U	SMPL
J4830-14 - 1	090618-DBRI-E-U-S	210	-0.0290	-245.1		PPB	11 Sep 2018 13:13:25	U	SMPL
J4830-14DUP - 1	090618-DBRI-E-U-SDUP	320	0.0035	-245.1		PPB	11 Sep 2018 13:15:40	U	SMPL
CCV29 - 1	CCV29	17499	5.0725	-245.1		PPB	11 Sep 2018 13:17:56	U	SMPL
CCB29 - 1	CCB29	-8	-0.0933	-245.1		PPB	11 Sep 2018 13:20:13	U	SMPL
J4830-14MS - 1	090618-DBRI-E-U-SMS	15101	4.3649	-245.1		PPB	11 Sep 2018 13:22:31	U	SMPL
J4830-14MSD - 1	090618-DBRI-E-U-SMSD	15430	4.4620	-245.1		PPB	11 Sep 2018 13:24:48	U	SMPL
J4830-14LX5 - 1		164	-0.0426	-245.1		PPB	11 Sep 2018 13:27:04	U	SMPL
J4830-14A - 1		17119	4.9604	-245.1		PPB	11 Sep 2018 13:29:22	U	SMPL
CCV30 - 1	CCV30	17310	5.0168	-245.1		PPB	11 Sep 2018 13:40:50	U	SMPL
CCB30 - 1	CCB30	140	-0.0497	-245.1		PPB	11 Sep 2018 13:43:06	U	SMPL