



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB97839

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.6	4.01	09/11/2018	12:20
2	CAL2	1	Water			21.8	7.01	09/11/2018	12:22
3	CAL3	1	Water			21.9	10.04	09/11/2018	12:24
4	ICV	1	Water			21.7	7.00	09/11/2018	12:26
5	CCV1	1	Water			21.9	2.00	09/11/2018	12:28
6	J4847-01	1	Solid	20.02	20	22.3	5.16	09/11/2018	12:30
7	J4847-01DUP	1	Solid	20.03	20	22.4	5.17	09/11/2018	12:32
8	CCV2	1	Water			21.8	12.00	09/11/2018	12:34

WORKLIST(Hardcopy Internal Chain)

LB 97839

WorkList Name : pH091118

WorkList ID : 116589

Date : 9/11/2018 12:12:49 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/10/2018	Solid	J4847-01	pH	Cool 4 deg C	TULL02	I52	FK-PS-01-040	09/10/2018	9045D

Date/Time 09/11/18 12:15
 Received by: SP
 Relinquished by: CP

Date/Time 09/11/18 1:00
 Received by: CP
 Relinquished by: SP