

7471B


LB97840

INSTRUMENT ID: CV1

Linear

 μ Abs.:

44771



A= 0.0000e+000

B= 2.2728e-004

C= -1.9662e-002

Rho= 0.9995811

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.032	0.032	226	0.000	226				
0.2	0.200	0.221	0.021	1061	0.0 %	1061				
2.5	2.500	2.538	0.038	11253	0.0 %	11253				
5.0	5.000	4.950	-0.050	21864	0.0 %	21864				
7.5	7.500	7.303	-0.197	32220	0.0 %	32220				
10.0	10.000	10.156	0.156	44771	0.0 %	44771				

LB97840 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 11 Sep 2018 14:28:38

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	226	-	0.00007471B	PPB		11 Sep 2018 14:31:16	S	Std
0.2 - 1	SO.2	1061	-	0.20007471B	PPB		11 Sep 2018 14:33:34	S	Std
2.5 - 1	SO.5	11253	-	2.50007471B	PPB		11 Sep 2018 14:35:52	S	Std
5.0 - 1	SO.5	21864	-	5.00007471B	PPB		11 Sep 2018 14:38:09	S	Std
7.5 - 1	SO.5	32220	-	7.50007471B	PPB		11 Sep 2018 14:40:27	S	Std
10.0 - 1	SO.5	44771	-	10.00007471B	PPB		11 Sep 2018 14:42:44	S	Std
ICV58 - 1	ICV58	17918	4.0528	-7471B	PPB		11 Sep 2018 14:45:49	U	SMPL
ICB58 - 1	ICB58	63	-0.0053	-7471B	PPB		11 Sep 2018 14:48:04	U	SMPL
CCV31 - 1	CCV31	22313	5.0517	-7471B	PPB		11 Sep 2018 14:50:21	U	SMPL
CCB31 - 1	CCB31	21	-0.0149	-7471B	PPB		11 Sep 2018 14:52:37	U	SMPL
CRA - 1	CRA	986	0.2044	-7471B	PPB		11 Sep 2018 14:54:54	U	SMPL
HighStd - 1	HighStd	44846	10.1730	-7471B	PPB		11 Sep 2018 14:57:10	U	SMPL
ChkStd - 1	ChkStd	30745	6.9681	-7471B	PPB		11 Sep 2018 14:59:26	U	SMPL
PB112750BL - 1	PBS	-111	-0.0449	-7471B	PPB		11 Sep 2018 15:01:44	U	SMPL
PB112750BS - 1	LCSS	14872	3.3605	-7471B	PPB		11 Sep 2018 15:04:03	U	SMPL
J4724-18 - 1	EP1-SUT-4USTS-SB	12336	2.7841	-7471B	PPB		11 Sep 2018 15:06:20	U	SMPL
J4838-07 - 1	SUP-1A-C-090718-1	3669	0.8142	-7471B	PPB		11 Sep 2018 15:08:38	U	SMPL
J4838-08 - 1	SUP-1B-C-090718-1	2190	0.4781	-7471B	PPB		11 Sep 2018 15:10:56	U	SMPL
J4842-01 - 1	SG-SOIL	1424	0.3040	-7471B	PPB		11 Sep 2018 15:13:14	U	SMPL
CCV32 - 1	CCV32	22268	5.0414	-7471B	PPB		11 Sep 2018 15:15:31	U	SMPL
CCB32 - 1	CCB32	53	-0.0076	-7471B	PPB		11 Sep 2018 15:17:47	U	SMPL
J4847-01 - 1	FK-PS-01-040	1214	0.2563	-7471B	PPB		11 Sep 2018 15:20:02	U	SMPL
J4847-01DUP - 1	FK-PS-01-040DUP	1242	0.2626	-7471B	PPB		11 Sep 2018 15:22:18	U	SMPL
J4847-01MS - 1	FK-PS-01-040MS	16830	3.8055	-7471B	PPB		11 Sep 2018 15:24:35	U	SMPL
J4847-01MSD - 1	FK-PS-01-040MSD	18265	4.1316	-7471B	PPB		11 Sep 2018 15:26:50	U	SMPL
J4847-01LX5 - 1		227	0.0319	-7471B	PPB		11 Sep 2018 15:29:07	U	SMPL
J4847-01A - 1		18660	4.2214	-7471B	PPB		11 Sep 2018 15:31:25	U	SMPL
CCV33 - 1	CCV33	22162	5.0173	-7471B	PPB		11 Sep 2018 15:40:35	U	SMPL
CCB33 - 1	CCB33	93	0.0015	-7471B	PPB		11 Sep 2018 15:42:51	U	SMPL