



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB97841

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.6	4.01	09/11/2018	12:20
2	CAL2	1	Water			21.8	7.01	09/11/2018	12:22
3	CAL3	1	Water			21.9	10.04	09/11/2018	12:24
4	ICV	1	Water			21.7	7.00	09/11/2018	12:26
5	CCV1	1	Water			21.7	2.00	09/11/2018	14:05
6	J4854-01	1	Water			18.6	5.27	09/11/2018	14:07
7	J4854-01DUP	1	Water			18.7	5.28	09/11/2018	14:09
8	CCV2	1	Water			21.9	12.00	09/11/2018	14:11

WORKLIST(Hardcopy Internal Chain)

LB97841

WorkList Name : pH091118-1

WorkList ID : 116591

Date : 9/11/2018 1:55:10 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/11/2018	Water	J4854-01 H	pH	Cool 4 deg C	MAJO01	I22	DRAIN-WATER-TANK-1	09/11/2018	9040C

Date/Time 09/11/18 2:00

Received by: SP

Relinquished by: CP

Date/Time 09/11/18 2:15

Received by: CP

Relinquished by: SP