



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB97844
Analysis Date: 09/11/2018
BalanceID: WC SC-3
OvenID: EX OVEN-2

ANALYST: JIGNESH
REVIEWED BY: apatel
Extraction Date: 09/11/2018
Extraction IN Time: 14:00
Extraction OUT Time: 14:35

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB97844BL	LB97844BL	WATER	2	1000	100	3.0658	3.0658	3.02	3.0658	3.0658	0.0000	0
2	LB97844BS	LB97844BS	WATER	2	1000	100	3.1547	3.1547	3.01	3.1732	3.1732	0.0185	18.5
3	LB97844BSD	LB97844BSD	WATER	2	1000	100	2.8803	2.8803	3.03	2.8991	2.8991	0.0188	18.8
4	J4854-01	DRAIN-WATER-TANK-1	WATER	2	1000	100	3.1327	3.1327	3.04	3.1369	3.1369	0.0042	4.2
5	J4854-01MS	DRAIN-WATER-TANK-1	WATER	2	1000	100	2.9792	2.9792	3.05	3.0035	3.0035	0.0243	24.3



QC Batch# LB97844 Test: TPH Analysis Date: 09/11/2018

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2408
pH Paper 0-14	W2373
Sodium Sulfate	EP1876
1:1 HCL	WP64228
Silica Gel	W2199
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP64229
LCSWD	10.00 ML	WP64230
MS/MSD	10.00 ML	WP64229

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance:	<u>0.0021</u>	In OVEN TEMP :	<u>70</u>	Dessicator Time In :	<u>16:11</u>
1.0000 gram Balance:	<u>1.0005</u>	In Time:	<u>15:10</u>		
Bal Check Time:	<u>14:15</u>	Out OVEN TEMP:	<u>70</u>	Dessicator Time Out:	<u>16:35</u>
		Out Time:	<u>16:10</u>		

After Analysis

0.0020 gram Balance:	<u>0.0019</u>	In OVEN TEMP :	<u>71</u>	Dessicator Time In :	<u>17:31</u>
1.0000 gram Balance:	<u>1.0003</u>	In Time:	<u>17:00</u>		
Bal Check Time:	<u>18:30</u>	Out OVEN TEMP:	<u>71</u>	Dessicator Time Out:	<u>18:00</u>
		Out Time:	<u>17:30</u>		

LB 97844

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TPH J4854

WorkList ID : 116598

Date : 9/11/2018 1:50:10 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/18/2018	Water	J4854-01	TPH	Conc H2SO4 to	MAJO01	I22	DRAIN-WATER-TANK-1	09/11/2018	1664A

E, F

Date/Time 09-11-18 21:00 PM
 Received by: Jo
 Relinquished by: CO

Date/Time 09-11-18 3:35 PM
 Received by: CO
 Relinquished by: Jo