



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB97845

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.6	4.01	09/11/2018	17:10
2	CAL2	1	Water			21.8	7.01	09/11/2018	17:12
3	CAL3	1	Water			21.7	10.04	09/11/2018	17:14
4	ICV	1	Water			21.8	7.00	09/11/2018	17:16
5	CCV1	1	Water			21.9	2.00	09/11/2018	17:18
6	J4780-02	1	Solid	20.02	20	23.3	7.82	09/11/2018	17:20
7	J4780-02DUP	1	Solid	20.03	20	23.4	7.83	09/11/2018	17:22
8	J4853-04	1	Solid	20.02	20	22.4	8.05	09/11/2018	17:24
9	J4856-04	1	Solid	20.04	20	22.9	9.02	09/11/2018	17:26
10	J4856-11	1	Solid	20.03	20	23.0	6.79	09/11/2018	17:28
11	CCV2	1	Water			21.8	12.00	09/11/2018	17:30

WORKLIST(Hardcopy Internal Chain)

LB 97845

WorkList Name : corrosivity091118

WorkList ID : 116593

Date : 9/11/2018 3:11:18 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/11/2018	Solid	J4780-02	Corrosivity	Cool 4 deg C	PSEG05		TR-04-091118	09/11/2018	9045D
09/11/2018	Solid	J4853-04	Corrosivity	Cool 4 deg C	PSEG01	I31	CS-SOIL	09/11/2018	9045D
09/11/2018	Solid	J4856-04	Corrosivity	Cool 4 deg C	PSEG01	I21	STONE-PILE	09/11/2018	9045D
09/11/2018	Solid	J4856-11	Corrosivity	Cool 4 deg C	PSEG01	I21	TP-1-3	09/11/2018	9045D

Date/Time 09/11/18 5:00
 Received by: SP
 Relinquished by: CP

Date/Time 09/11/18 5:35
 Received by: CP
 Relinquished by: SP