

245.1


LB 97851

INSTRUMENT ID: CV4

Linear

 μ Abs.:

35513



A= 0.0000e+000

B= 2.8161e-004

C= 1.6232e-002

Rho= 0.9999228

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.021	0.021	18	0.000	18				
0.05	0.050	0.062	0.012	161	0.0 %	161				
0.2	0.200	0.199	-0.001	650	0.0 %	650				
2.5	2.500	2.425	-0.075	8554	0.0 %	8554				
5.0	5.000	5.076	0.076	17969	0.0 %	17969				
7.5	7.500	7.449	-0.051	26395	0.0 %	26395				
10.0	10.000	10.017	0.017	35513	0.0 %	35513				

LB97851 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 12 Sep 2018 10:46:16

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	18	-	0.0000	245.1	PPB	12 Sep 2018 10:47:55	S	Std
0.05 - 1	50 05	161	-	0.0500	245.1	PPB	12 Sep 2018 10:50:12	S	Std
0.2 - 1	50 2	650	-	0.2000	245.1	PPB	12 Sep 2018 10:52:30	S	Std
2.5 - 1	52.5	8554	-	2.5000	245.1	PPB	12 Sep 2018 10:54:47	S	Std
5.0 - 1	55	17969	-	5.0000	245.1	PPB	12 Sep 2018 10:57:05	S	Std
7.5 - 1	57.5	26395	-	7.5000	245.1	PPB	12 Sep 2018 10:59:21	S	Std
10.0 - 1	510	35513	-	10.0000	245.1	PPB	12 Sep 2018 11:04:16	S	Std
ICV60 - 1	ICV60	13665	3.8644	-245.1		PPB	12 Sep 2018 11:07:17	U	SMPL
ICB60 - 1	ICB60	-62	-0.0012	-245.1		PPB	12 Sep 2018 11:09:33	U	SMPL
CCV37 - 1	CCV37	17911	5.0601	-245.1		PPB	12 Sep 2018 11:11:50	U	SMPL
CCB37 - 1	CCB37	-299	-0.0680	-245.1		PPB	12 Sep 2018 11:14:06	U	SMPL
CRA - 1	CRA	171	0.0644	-245.1		PPB	12 Sep 2018 11:16:22	U	SMPL
HighStd - 1	HighStd	35148	9.9142	-245.1		PPB	12 Sep 2018 11:24:26	U	SMPL
ChkStd - 1	ChkStd	24207	6.8331	-245.1		PPB	12 Sep 2018 11:26:42	U	SMPL
PB112770BL - 1	PBW	-314	-0.0722	-245.1		PPB	12 Sep 2018 11:29:00	U	SMPL
PB112770BS - 1	LCSW	13466	3.8084	-245.1		PPB	12 Sep 2018 11:31:19	U	SMPL
J4841-01 - 1	090718-DBRI-F-D-B	-111	-0.0150	-245.1		PPB	12 Sep 2018 11:33:36	U	SMPL
J4841-02 - 1	090718-DBRI-F-D-M	-2	0.0157	-245.1		PPB	12 Sep 2018 11:35:54	U	SMPL
J4841-03 - 1	090718-DBRI-F-D-S	21	0.0221	-245.1		PPB	12 Sep 2018 11:38:12	U	SMPL
J4841-04 - 1	090718-DBRI-F-U-B	110	0.0472	-245.1		PPB	12 Sep 2018 11:40:30	U	SMPL
J4841-05 - 1	090718-DBRI-F-U-M	-31	0.0075	-245.1		PPB	12 Sep 2018 11:42:46	U	SMPL
CCV38 - 1	CCV38	17776	5.0221	-245.1		PPB	12 Sep 2018 11:45:01	U	SMPL
CCB38 - 1	CCB38	-204	-0.0412	-245.1		PPB	12 Sep 2018 11:47:17	U	SMPL
J4841-06 - 1	090718-DBRI-F-U-S	-20	0.0106	-245.1		PPB	12 Sep 2018 11:49:35	U	SMPL
J4841-07 - 1	090718-SIDI-F-D-B	65	0.0345	-245.1		PPB	12 Sep 2018 11:51:52	U	SMPL
J4841-08 - 1	090718-SIDI-F-D-M	100	0.0444	-245.1		PPB	12 Sep 2018 11:54:08	U	SMPL
J4841-09 - 1	090718-SIDI-F-D-S	19	0.0216	-245.1		PPB	12 Sep 2018 11:56:24	U	SMPL
J4841-10 - 1	090718-SIDI-F-U-B	-43	0.0041	-245.1		PPB	12 Sep 2018 11:58:40	U	SMPL
J4841-11 - 1	090718-SIDI-F-U-M	62	0.0337	-245.1		PPB	12 Sep 2018 12:00:58	U	SMPL
J4841-12 - 1	090718-SIDI-F-U-S	33	0.0255	-245.1		PPB	12 Sep 2018 12:03:15	U	SMPL
J4841-12DUP - 1	090718-SIDI-F-U-SDUP	-36	0.0061	-245.1		PPB	12 Sep 2018 12:05:32	U	SMPL
J4841-12MS - 1	090718-SIDI-F-U-SMS	15093	4.2665	-245.1		PPB	12 Sep 2018 12:07:50	U	SMPL
J4841-12MSD - 1	090718-SIDI-F-U-SMSD	16022	4.5282	-245.1		PPB	12 Sep 2018 12:10:06	U	SMPL
CCV39 - 1	CCV39	17359	4.9047	-245.1		PPB	12 Sep 2018 12:12:23	U	SMPL
CCB39 - 1	CCB39	-242	-0.0519	-245.1		PPB	12 Sep 2018 12:14:40	U	SMPL
J4841-12LX5 - 1		-23	0.0098	-245.1		PPB	12 Sep 2018 12:16:59	U	SMPL
J4841-12A - 1		16886	4.7715	-245.1		PPB	12 Sep 2018 12:19:15	U	SMPL
CCV40 - 1	CCV40	17166	4.8503	-245.1		PPB	12 Sep 2018 12:21:32	U	SMPL
CCV40 - 1	CCV40	207	0.0745	-245.1		PPB	12 Sep 2018 12:23:48	U	SMPL