



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB97866

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.1	4.01	09/12/2018	16:40
2	CAL2	1	Water			21.9	7.01	09/12/2018	16:42
3	CAL3	1	Water			22.0	10.04	09/12/2018	16:44
4	ICV	1	Water			22.1	7.00	09/12/2018	16:46
5	CCV1	1	Water			21.8	2.00	09/12/2018	17:05
6	J4771-02	1	Solid	20.02	20	23.8	10.54	09/12/2018	17:07
7	J4771-02DUP	1	Solid	20.04	20	23.9	10.55	09/12/2018	17:09
8	J4873-02	1	Solid	20.02	20	23.3	10.38	09/12/2018	17:11
9	J4873-04	1	Solid	20.03	20	23.2	10.77	09/12/2018	17:13
10	J4873-06	1	Solid	20.02	20	23.0	9.10	09/12/2018	17:15
11	J4883-02	1	Solid	20.02	20	23.9	9.01	09/12/2018	17:17
12	CCV2	1	Water			21.9	12.00	09/12/2018	17:19

WORKLIST(Hardcopy Internal Chain)

LB97866

WorkList Name : corrosivity091218

WorkList ID : 116631

Date : 9/12/2018 3:30:41 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/12/2018	Solid	J4771-02	Corrosivity	Cool 4 deg C	PSEG05	I31	NB-08-091218	09/12/2018	9045D
09/12/2018	Solid	J4873-02	Corrosivity	Cool 4 deg C	PSEG01	I31	NF-CON-1	09/12/2018	9045D
09/12/2018	Solid	J4873-04	Corrosivity	Cool 4 deg C	PSEG01	I31	NF-CON-2	09/12/2018	9045D
09/12/2018	Solid	J4873-06	Corrosivity	Cool 4 deg C	PSEG01	I31	NF-STONE	09/12/2018	9045D
09/12/2018	Solid	J4883-02	Corrosivity	Cool 4 deg C	PSEG01	I31	292	09/12/2018	9045D

Date/Time 09/12/18 4:30

Received by: SP

Relinquished by: CP

Date/Time 09/12/18 5:25

Received by: CP

Relinquished by: SP