



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB97867

Slope : 99.1

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.1	4.01	09/12/2018	16:40
2	CAL2	1	Water			21.9	7.01	09/12/2018	16:42
3	CAL3	1	Water			22.0	10.04	09/12/2018	16:44
4	ICV	1	Water			22.1	7.00	09/12/2018	16:46
5	CCV1	1	Water			21.9	2.00	09/12/2018	16:48
6	J4882-02	1	Water			22.0	7.25	09/12/2018	16:50
7	J4882-02DUP	1	Water			22.1	7.26	09/12/2018	16:52
8	J4883-04	1	Water			23.1	6.99	09/12/2018	16:54
9	CCV2	1	Water			21.8	12.00	09/12/2018	16:56

LB a7867

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity091218-1

WorkList ID : 116632

Date : 9/12/2018 3:30:51 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/12/2018	Water	J4882-02	Corrosivity	Cool 4 deg C	PSEG01	I31	28570	09/12/2018	9040C
09/12/2018	Water	J4883-04	Corrosivity	Cool 4 deg C	PSEG01	I31	SV30518L	09/12/2018	9040C

Date/Time 09/12/18 4:30
 Received by: Sf
 Relinquished by: Sf

Date/Time 09/12/18 5:00
 Received by: Sf
 Relinquished by: Sf