



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB97878

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.0	80.6	80.6	09/13/2018	09:20
2	J4882-02		1	7	>100.0	>212.0	>212.0	09/13/2018	09:45
3	J4882-02DUP		1	7	>100.0	>212.0	>212.0	09/13/2018	10:15
4	J4883-04		1	9	>100.0	>212.0	>212.0	09/13/2018	10:42

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp91318

WorkList ID : 116658

Date : 9/13/2018 8:54:01 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/17/2018	Water	J4882-02	Flash Point	Cool 4 deg C	PSEG01	I31	28570	09/12/2018	1010A
09/17/2018	Water	J4883-04	Flash Point	Cool 4 deg C	PSEG01	I31	SV30518L	09/12/2018	1010A

LB 97878

Date/Time 09/13/18 8:55AM
 Received by: KS
 Relinquished by: CP

Date/Time 09/13/18 10:45AM
 Received by: CP
 Relinquished by: KS