



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB97879

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.1	4.01	09/12/2018	16:40
2	CAL2	1	Water			21.9	7.01	09/12/2018	16:42
3	CAL3	1	Water			22.0	10.04	09/12/2018	16:44
4	ICV	1	Water			22.1	7.00	09/12/2018	16:46
5	CCV1	1	Water			21.9	2.00	09/12/2018	17:20
6	J4866-01	1	Solid	20.02	20	23.4	9.05	09/12/2018	17:22
7	J4866-01DUP	1	Solid	20.03	20	23.5	9.06	09/12/2018	17:24
8	J4866-02	1	Solid	20.02	20	23.1	9.02	09/12/2018	17:26
9	CCV2	1	Water			21.9	12.00	09/12/2018	17:28

WORKLIST(Hardcopy Internal Chain)

LB 97879

WorkList Name : pH091218

WorkList ID : 116633

Date : 9/12/2018 3:31:01 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/11/2018	Solid	J4866-01	pH	Cool 4 deg C	SHAW22	I41	SP-545-180911	09/11/2018	9045D
09/11/2018	Solid	J4866-02	pH	Cool 4 deg C	SHAW22	I41	DPT-SB-E1-180911	09/11/2018	9045D

Date/Time 09/12/18 4:30

Received by: SP

Relinquished by: CP

Date/Time 09/12/18 5:30

Received by: CP

Relinquished by: SP