

245.1


LB97883

INSTRUMENT ID: CV1

Linear

 μ Abs.:

34423



A= 0.0000e+000

B= 2.8918e-004

C= -1.5431e-002

Rho= 0.9999425

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.005	0.005	70	0.000	70				
0.05	0.050	0.046	-0.004	212	0.0 %	212				
0.2	0.200	0.191	-0.009	713	0.0 %	713				
2.5	2.500	2.463	-0.037	8571	0.0 %	8571				
5.0	5.000	5.042	0.042	17489	0.0 %	17489				
7.5	7.500	7.564	0.064	26211	0.0 %	26211				
10.0	10.000	9.939	-0.061	34423	0.0 %	34423				

LB97883 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 13 Sep 2018 12:14:08

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	70	-	0.0000	245.1	PPB	13 Sep 2018 12:17:29	S	Std
0.05 - 1	50.05	212	-	0.0500	245.1	PPB	13 Sep 2018 12:25:15	S	Std
0.2 - 1	50.2	713	-	0.2000	245.1	PPB	13 Sep 2018 12:27:33	S	Std
2.5 - 1	52.5	8571	-	2.5000	245.1	PPB	13 Sep 2018 12:29:50	S	Std
5.0 - 1	55	17489	-	5.0000	245.1	PPB	13 Sep 2018 12:32:07	S	Std
7.5 - 1	57.5	26211	-	7.5000	245.1	PPB	13 Sep 2018 12:34:26	S	Std
10.0 - 1	510	34423	-	10.0000	245.1	PPB	13 Sep 2018 12:36:43	S	Std
ICV63 - 1	ICV63	13752	3.9614	-245.1		PPB	13 Sep 2018 12:39:33	U	SMPL
ICB63 - 1	ICB63	-187	-0.0695	-245.1		PPB	13 Sep 2018 12:41:50	U	SMPL
CCV47 - 1	CCV47	17362	5.0053	-245.1		PPB	13 Sep 2018 12:44:07	U	SMPL
CCB47 - 1	CCB47	-213	-0.0770	-245.1		PPB	13 Sep 2018 12:46:24	U	SMPL
CRA - 1	CRA	190	0.0395	-245.1		PPB	13 Sep 2018 12:48:42	U	SMPL
HighStd - 1	HighStd	34402	9.9330	-245.1		PPB	13 Sep 2018 12:50:57	U	SMPL
ChkStd - 1	ChkStd	24370	7.0319	-245.1		PPB	13 Sep 2018 12:53:13	U	SMPL
PB112793BL - 1	PBW	-35	-0.0256	-245.1		PPB	13 Sep 2018 12:58:10	U	SMPL
PB112793BS - 1	LCSW	14091	4.0594	-245.1		PPB	13 Sep 2018 13:00:27	U	SMPL
J4790-01 - 1	EFFLUENT-DAY1-METALS	106	0.0152	-245.1		PPB	13 Sep 2018 13:07:36	U	SMPL
J4790-01DUP - 1	EFFLUENT-DAY1-METALS DUP	113	0.0172	-245.1		PPB	13 Sep 2018 13:09:53	U	SMPL
J4790-01MS - 1	EFFLUENT-DAY1-METALS MS	12316	3.5461	-245.1		PPB	13 Sep 2018 13:12:10	U	SMPL
J4790-01MSD - 1	EFFLUENT-DAY1-METALS MSD	12358	3.5583	-245.1		PPB	13 Sep 2018 13:14:28	U	SMPL
J4811-02 - 1	090518-DBRI-E-D-M	-92	-0.0420	-245.1		PPB	13 Sep 2018 13:16:44	U	SMPL
CCV48 - 1	CCV48	16322	4.7046	-245.1		PPB	13 Sep 2018 13:22:11	U	SMPL
CCB48 - 1	CCB48	-161	-0.0620	-245.1		PPB	13 Sep 2018 13:24:26	U	SMPL
J4811-03 - 1	090518-DBRI-E-D-S	-57	-0.0319	-245.1		PPB	13 Sep 2018 13:26:45	U	SMPL
J4811-04 - 1	090518-DBRI-E-U-B	33	-0.0059	-245.1		PPB	13 Sep 2018 13:29:00	U	SMPL
J4811-05 - 1	090518-DBRI-E-U-M	-64	-0.0339	-245.1		PPB	13 Sep 2018 13:31:16	U	SMPL
J4811-06 - 1	090518-DBRI-E-U-S	85	0.0091	-245.1		PPB	13 Sep 2018 13:33:32	U	SMPL
J4811-07 - 1	090518-SIDI-E-D-B	-3	-0.0163	-245.1		PPB	13 Sep 2018 13:35:49	U	SMPL
J4811-08 - 1	090518-SIDI-E-D-S	-37	-0.0261	-245.1		PPB	13 Sep 2018 13:38:06	U	SMPL
J4811-09 - 1	090518-SIDI-E-U-B	32	-0.0062	-245.1		PPB	13 Sep 2018 13:40:23	U	SMPL
J4811-10 - 1	090518-SIDI-E-U-S	15	-0.0111	-245.1		PPB	13 Sep 2018 13:42:40	U	SMPL
J4811-11 - 1	090518-TIPC-E-D-M	37	-0.0047	-245.1		PPB	13 Sep 2018 13:44:58	U	SMPL
J4811-12 - 1	090518-TIPC-E-U-M	25	-0.0082	-245.1		PPB	13 Sep 2018 13:47:15	U	SMPL
CCV49 - 1	CCV49	16389	4.7240	-245.1		PPB	13 Sep 2018 13:49:31	U	SMPL
CCB49 - 1	CCB49	-144	-0.0571	-245.1		PPB	13 Sep 2018 13:51:46	U	SMPL
J4790-01LX5 - 1		89	0.0103	-245.1		PPB	13 Sep 2018 13:54:05	U	SMPL
J4790-01A - 1		13090	3.7700	-245.1		PPB	13 Sep 2018 13:56:21	U	SMPL
CCV50 - 1	CCV50	16606	4.7867	-245.1		PPB	13 Sep 2018 13:58:38	U	SMPL
CCB50 - 1	CCB50	-14	-0.0195	-245.1		PPB	13 Sep 2018 14:03:47	U	SMPL