

7471B


LB97892

INSTRUMENT ID: CV7

Linear

 μ Abs.:

40614



A= 0.0000e+000

B= 2.4317e-004

C= -1.2532e-002

Rho= 0.9995215

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.001	-0.001	46	0.000	46				
0.2	0.200	0.181	-0.019	795	0.0 %	795				
2.5	2.500	2.522	0.022	10425	0.0 %	10425				
5.0	5.000	4.910	-0.090	20243	0.0 %	20243				
7.5	7.500	7.725	0.225	31819	0.0 %	31819				
10.0	10.000	9.863	-0.137	40614	0.0 %	40614				

LB97892 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 13 Sep 2018 15:51:17

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	46	-	0.0000	7471B	PPB	13 Sep 2018 15:56:09	S	Std
0.2 - 1	50.2	795	-	0.2000	7471B	PPB	13 Sep 2018 15:58:26	S	Std
2.5 - 1	52.5	10425	-	2.5000	7471B	PPB	13 Sep 2018 16:00:44	S	Std
5.0 - 1	55	20243	-	5.0000	7471B	PPB	13 Sep 2018 16:03:01	S	Std
7.5 - 1	57.5	31819	-	7.5000	7471B	PPB	13 Sep 2018 16:07:53	S	Std
10.0 - 1	511	40614	-	10.0000	7471B	PPB	13 Sep 2018 16:10:09	S	Std
ICV64 - 1	ICV64	15132	3.6671	-	7471B	PPB	13 Sep 2018 16:12:51	U	SMPL
ICB64 - 1	ICB64	-124	-0.0427	-	7471B	PPB	13 Sep 2018 16:15:09	U	SMPL
CCV51 - 1	CCV51	20524	4.9782	-	7471B	PPB	13 Sep 2018 16:17:28	U	SMPL
CCB51 - 1	CCB51	-172	-0.0544	-	7471B	PPB	13 Sep 2018 16:19:44	U	SMPL
CRA - 1	CRA	829	0.1891	-	7471B	PPB	13 Sep 2018 16:22:02	U	SMPL
HighStd - 1	HighStd	41165	9.9974	-	7471B	PPB	13 Sep 2018 16:24:18	U	SMPL
ChkStd - 1	ChkStd	29802	7.2343	-	7471B	PPB	13 Sep 2018 16:26:34	U	SMPL
PB112814BL - 1	PBS	-207	-0.0629	-	7471B	PPB	13 Sep 2018 16:28:54	U	SMPL
PB112814BS - 1	LCSS	14929	3.6177	-	7471B	PPB	13 Sep 2018 16:31:14	U	SMPL
J4771-01 - 1	NB-08-091218	2332	0.5545	-	7471B	PPB	13 Sep 2018 16:38:30	U	SMPL
J4771-01DUP - 1	NB-08-091218DUP	2080	0.4933	-	7471B	PPB	13 Sep 2018 16:40:47	U	SMPL
J4771-01MS - 1	NB-08-091218MS	20503	4.9731	-	7471B	PPB	13 Sep 2018 16:43:04	U	SMPL
J4771-01MSD - 1	NB-08-091218MSD	19496	4.7282	-	7471B	PPB	13 Sep 2018 16:45:22	U	SMPL
J4873-01 - 1	NF-CON-1	293	0.0587	-	7471B	PPB	13 Sep 2018 16:47:39	U	SMPL
CCV52 - 1	CCV52	20983	5.0898	-	7471B	PPB	13 Sep 2018 16:49:55	U	SMPL
CCB52 - 1	CCB52	-99	-0.0366	-	7471B	PPB	13 Sep 2018 16:52:11	U	SMPL
J4873-03 - 1	NF-CON-2	282	0.0560	-	7471B	PPB	13 Sep 2018 16:54:27	U	SMPL
J4873-05 - 1	NF-STONE	105	0.0130	-	7471B	PPB	13 Sep 2018 16:56:43	U	SMPL
J4883-01 - 1	292	13441	3.2559	-	7471B	PPB	13 Sep 2018 16:58:59	U	SMPL
J4771-01LX5 - 1		364	0.0760	-	7471B	PPB	13 Sep 2018 17:01:16	U	SMPL
J4771-01A - 1		21222	5.1479	-	7471B	PPB	13 Sep 2018 17:03:32	U	SMPL
CCV53 - 1	CCV53	21028	5.1008	-	7471B	PPB	13 Sep 2018 17:05:50	U	SMPL
CCB53 - 1	CCB53	-157	-0.0507	-	7471B	PPB	13 Sep 2018 17:08:08	U	SMPL