



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB97897

Slope : 98.9

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.7	4.01	09/13/2018	16:50
2	CAL2	1	Water			22.5	7.01	09/13/2018	16:52
3	CAL3	1	Water			22.8	10.04	09/13/2018	16:54
4	ICV	1	Water			22.5	7.00	09/13/2018	16:56
5	CCV1	1	Water			22.6	2.00	09/13/2018	16:58
6	J4906-04	1	Solid	20.02	20	24.2	6.06	09/13/2018	17:00
7	J4906-04DUP	1	Solid	20.03	20	24.3	6.07	09/13/2018	17:02
8	J4906-08	1	Solid	20.02	20	24.3	6.67	09/13/2018	17:04
9	J4906-12	1	Solid	20.02	20	24.1	6.58	09/13/2018	17:06
10	J4906-16	1	Solid	20.04	20	24.2	5.64	09/13/2018	17:08
11	J4906-20	1	Solid	20.02	20	24.0	6.43	09/13/2018	17:10
12	J4906-24	1	Solid	20.02	20	23.7	5.66	09/13/2018	17:12
13	J4906-28	1	Solid	20.03	20	23.6	6.24	09/13/2018	17:14
14	J4906-32	1	Solid	20.02	20	23.5	6.63	09/13/2018	17:16
15	CCV2	1	Water			22.6	12.00	09/13/2018	17:18

WORKLIST(Hardcopy Internal Chain)

LB 97897

WorkList Name : corrosivity091318

WorkList ID : 116689

Date : 9/13/2018 4:14:10 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/13/2018	Solid	J4906-04	Corrosivity	Cool 4 deg C	PSEG01	I31	SP-PIT-3	09/13/2018	9045D
09/13/2018	Solid	J4906-08	Corrosivity	Cool 4 deg C	PSEG01	I31	SP-PIT-1	09/13/2018	9045D
09/13/2018	Solid	J4906-12	Corrosivity	Cool 4 deg C	PSEG01	I31	SP-PIT-7	09/13/2018	9045D
09/13/2018	Solid	J4906-16	Corrosivity	Cool 4 deg C	PSEG01	I31	SP-PIT-8	09/13/2018	9045D
09/13/2018	Solid	J4906-20	Corrosivity	Cool 4 deg C	PSEG01	I31	SP-PIT-9	09/13/2018	9045D
09/13/2018	Solid	J4906-24	Corrosivity	Cool 4 deg C	PSEG01	I31	SP-PIT-10	09/13/2018	9045D
09/13/2018	Solid	J4906-28	Corrosivity	Cool 4 deg C	PSEG01	I31	SP-PIT-4	09/13/2018	9045D
09/13/2018	Solid	J4906-32	Corrosivity	Cool 4 deg C	PSEG01	I31	SP-PIT-2	09/13/2018	9045D

Date/Time 09/13/18 4:35
 Received by: SP
 Relinquished by: CP

Date/Time 09/13/18 5:30
 Received by: CP
 Relinquished by: SP