



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB97909

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP67780
Turbidity std, 40NTU	WP67775
turbidity std 5 NTU	WP67778
turbidity std 20 NTU	WP67776
Turbidity std 1 NTU	WP67779
turbidity std 0 NTU	WP67781
turbidity std 10 NTU	WP67777
10 NTU Standard 500 ml	W2297

Intercept: 0.3464

Slope: 0.9055

Regression: 0.9998

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.297	09/14/2018	09:45
2	CAL2	1	1	0.978	09/14/2018	09:48
3	CAL3	5	1	4.760	09/14/2018	09:50
4	CAL4	10	1	9.860	09/14/2018	09:53
5	CAL5	20	1	18.600	09/14/2018	09:55
6	CAL6	40	1	36.400	09/14/2018	10:00



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB97909

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.020	9.579	09/14/2018	10:04
2	ICB		1	0.290	-0.062	09/14/2018	10:06
3	CCV1	10	1	9.280	9.866	09/14/2018	10:10
4	CCB1		1	0.292	-0.060	09/14/2018	10:14
5	LB97909BL		1	0.296	-0.056	09/14/2018	10:17
6	J4889-01		1	20.400	22.146	09/14/2018	10:19
7	J4889-01DUP		1	22.400	24.355	09/14/2018	10:22
8	CCV2	10	1	9.360	9.954	09/14/2018	10:25
9	CCB2		1	0.287	-0.066	09/14/2018	10:28

LB94909

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tur091318

WorkList ID : 116670

Date : 9/13/2018 11:52:11 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/15/2018	Water	J4889-01	Turbidity	Cool 4 deg C	LIRO01	I32	UDS-ER22-01	09/13/2018	SM2130 B

Date/Time 9-14-18 9:20
 Received by: AB
 Relinquished by: CR

Date/Time 9-14-18 10:35
 Received by: CR
 Relinquished by: AB