


ISM02.4

LB97911
INSTRUMENT ID: CV1

Linear

 μ Abs.:

4560E



A= 0.0000e+000

B= 2.1950e-004

C= 2.5708e-002

Rho= 0.9999571

Accept=Accepted

Slope
y=1015.8x

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%ob
0.00	0.000	0.033	0.033	31	0.000	31					14
0.2	0.200	0.227	0.027	915	0.0 %	915					-2
2.5	2.500	2.451	-0.049	11051	0.0 %	11051					-1
5.0	5.000	4.962	-0.038	22487	0.0 %	22487					0
7.5	7.500	7.491	-0.009	34013	0.0 %	34013					0
10.0	10.000	10.037	0.037	45608	0.0 %	45608					0

103
9114118

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	31	-		0 ISM02.4	PPB	9/14/2018 11:07	Std
	0.2 S0.2	915	-		0.2 ISM02.4	PPB	9/14/2018 11:10	Std
	2.5 S2.5	11051	-		2.5 ISM02.4	PPB	9/14/2018 11:12	Std
	5 S5	22487	-		5 ISM02.4	PPB	9/14/2018 11:14	Std
	7.5 S7.5	34013	-		7.5 ISM02.4	PPB	9/14/2018 11:17	Std
	10 S10.0	45608	-		10 ISM02.4	PPB	9/14/2018 11:19	Std
ICV	ICV	17408	3.8467 -		ISM02.4	PPB	9/14/2018 11:22	SMPL
ICB	ICB	-65	0.0114 -		ISM02.4	PPB	9/14/2018 11:24	SMPL
CCV059	CCV059	22947	5.0625 -		ISM02.4	PPB	9/14/2018 11:26	SMPL
CCB059	CCB059	-151	-0.0074 -		ISM02.4	PPB	9/14/2018 11:28	SMPL
PB112826BL	PBS004	-16	0.0222 -		ISM02.4	PPB	9/14/2018 11:31	SMPL
J4861-01	MA41T2	13100	2.9011 -		ISM02.4	PPB	9/14/2018 11:33	SMPL
J4861-02	MA41T2D	13466	2.9815 -		ISM02.4	PPB	9/14/2018 11:40	SMPL
J4861-03	MA41T2S	22452	4.9539 -		ISM02.4	PPB	9/14/2018 11:43	SMPL
PB112827BL	PBS004	-303	-0.0408 -		ISM02.4	PPB	9/14/2018 11:45	SMPL
J4886-01	MH2TS2	747	0.1897 -		ISM02.4	PPB	9/14/2018 11:47	SMPL
J4886-02	MH2TS2D	721	0.184 -		ISM02.4	PPB	9/14/2018 11:50	SMPL
J4886-03	MH2TS2S	12456	2.7598 -		ISM02.4	PPB	9/14/2018 11:52	SMPL
CCV060	CCV060	23087	5.0932 -		ISM02.4	PPB	9/14/2018 11:54	SMPL
CCB060	CCB060	-122	-0.0011 -		ISM02.4	PPB	9/14/2018 11:56	SMPL