

LB 97917


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

34988



A= 0.0000e+000

B= 2.8379e-004

C= 2.5609e-002

Rho= 0.9999309

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.026	0.026	1	0.000	1				
0.2	0.200	0.196	-0.004	599	0.0 %	599				
2.5	2.500	2.445	-0.055	8527	0.0 %	8527				
5.0	5.000	5.004	0.004	17541	0.0 %	17541				
7.5	7.500	7.575	0.075	26601	0.0 %	26601				
10.0	10.000	9.955	-0.045	34988	0.0 %	34988				

LB97917 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 14 Sep 2018 13:01:20

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	1	-	0.0000	7470A	PPB	14 Sep 2018 13:37:41	S	Std
0.2 - 1	50.2	599	-	0.2000	7470A	PPB	14 Sep 2018 13:39:49	S	Std
2.5 - 1	52.5	8527	-	2.5000	7470A	PPB	14 Sep 2018 13:41:58	S	Std
5.0 - 1	55	17541	-	5.0000	7470A	PPB	14 Sep 2018 13:44:06	S	Std
7.5 - 1	57.5	26601	-	7.5000	7470A	PPB	14 Sep 2018 13:46:14	S	Std
10.0 - 1	510	34988	-	10.0000	7470A	PPB	14 Sep 2018 13:48:23	S	Std
ICV66 - 1	ICV66	13349	3.8139	-7470A	PPB	14 Sep 2018 13:50:54	U	SMPL	
ICB66 - 1	ICB66	-259	-0.0479	-7470A	PPB	14 Sep 2018 13:53:01	U	SMPL	
CCV61 - 1	CCV61	17713	5.0524	-7470A	PPB	14 Sep 2018 13:55:09	U	SMPL	
CCB61 - 1	CCB61	-256	-0.0470	-7470A	PPB	14 Sep 2018 13:57:15	U	SMPL	
CRA - 1	CRA	535	0.1774	-7470A	PPB	14 Sep 2018 13:59:25	U	SMPL	
HighStd - 1	HighStd	35281	10.0380	-7470A	PPB	14 Sep 2018 14:01:31	U	SMPL	
ChkStd - 1	ChkStd	24740	7.0465	-7470A	PPB	14 Sep 2018 14:03:38	U	SMPL	
PB112816BL - 1	PBW	-80	0.0029	-7470A	PPB	14 Sep 2018 14:08:18	U	SMPL	
PB112816BS - 1	LCSW	13259	3.7884	-7470A	PPB	14 Sep 2018 14:21:14	U	SMPL	
PB112816TB - 1	PB112797TB	-224	-0.0380	-7470A	PPB	14 Sep 2018 14:23:22	U	SMPL	
J4771-02 - 1	NB-08-091218	-69	0.0060	-7470A	PPB	14 Sep 2018 14:25:31	U	SMPL	
J4771-02DUP - 1	NB-08-091218DUP	-46	0.0126	-7470A	PPB	14 Sep 2018 14:27:40	U	SMPL	
J4771-02MS - 1	NB-08-091218MS	13404	3.8295	-7470A	PPB	14 Sep 2018 14:29:49	U	SMPL	
J4771-02MSD - 1	NB-08-091218MSD	12802	3.6587	-7470A	PPB	14 Sep 2018 14:31:55	U	SMPL	
CCV62 - 1	CCV62	17653	5.0353	-7470A	PPB	14 Sep 2018 14:34:02	U	SMPL	
CCB62 - 1	CCB62	-297	-0.0587	-7470A	PPB	14 Sep 2018 14:36:10	U	SMPL	
J4866-01 - 1	SP-545-180911	-85	0.0015	-7470A	PPB	14 Sep 2018 14:38:18	U	SMPL	
J4866-02 - 1	DPT-SB-E1-180911	-12	0.0222	-7470A	PPB	14 Sep 2018 14:40:25	U	SMPL	
J4873-02 - 1	NF-CON-1	-25	0.0185	-7470A	PPB	14 Sep 2018 14:42:32	U	SMPL	
J4873-04 - 1	NF-CON-2	16	0.0301	-7470A	PPB	14 Sep 2018 14:44:39	U	SMPL	
J4873-06 - 1	NF-STONE	-84	0.0018	-7470A	PPB	14 Sep 2018 14:46:46	U	SMPL	
J4883-02 - 1	292	366	0.1295	-7470A	PPB	14 Sep 2018 14:48:54	U	SMPL	
PB112817BL - 1	PBW	-10	0.0228	-7470A	PPB	14 Sep 2018 14:51:02	U	SMPL	
PB112817BS - 1	LCSW	13381	3.8230	-7470A	PPB	14 Sep 2018 14:53:10	U	SMPL	
PB112817TB - 1	PB112799TB	-205	-0.0326	-7470A	PPB	14 Sep 2018 14:55:19	U	SMPL	
J4882-02 - 1	28570	-26	0.0182	-7470A	PPB	14 Sep 2018 14:57:27	U	SMPL	
CCV63 - 1	CCV63	17318	4.9403	-7470A	PPB	14 Sep 2018 14:59:34	U	SMPL	
CCB63 - 1	CCB63	-219	-0.0365	-7470A	PPB	14 Sep 2018 15:01:40	U	SMPL	
J4882-02DUP - 1	28570DUP	-13	0.0219	-7470A	PPB	14 Sep 2018 15:03:48	U	SMPL	
J4882-02MS - 1	28570MS	14987	4.2788	-7470A	PPB	14 Sep 2018 15:05:55	U	SMPL	
J4882-02MSD - 1	28570MSD	13293	3.7980	-7470A	PPB	14 Sep 2018 15:08:02	U	SMPL	
J4883-04 - 1	SV30518L	-252	-0.0459	-7470A	PPB	14 Sep 2018 15:10:11	U	SMPL	
PB112818BL - 1	PBW	-58	0.0091	-7470A	PPB	14 Sep 2018 15:12:19	U	SMPL	
PB112818BS - 1	LCSW	12664	3.6195	-7470A	PPB	14 Sep 2018 15:14:27	U	SMPL	
J4882-01 - 1	28570	264	0.1005	-7470A	PPB	14 Sep 2018 15:25:19	U	SMPL	
J4882-01DUP - 1	28570DUP	215	0.0866	-7470A	PPB	14 Sep 2018 15:27:27	U	SMPL	
J4882-01MS - 1	28570MS	15563	4.4422	-7470A	PPB	14 Sep 2018 15:29:35	U	SMPL	
J4882-01MSD - 1	28570MSD	15167	4.3298	-7470A	PPB	14 Sep 2018 15:31:43	U	SMPL	
CCV64 - 1	CCV64	17465	4.9820	-7470A	PPB	14 Sep 2018 15:33:51	U	SMPL	
CCB64 - 1	CCB64	-259	-0.0479	-7470A	PPB	14 Sep 2018 15:35:59	U	SMPL	
J4883-03 - 1	SV30518L	54	0.0409	-7470A	PPB	14 Sep 2018 15:38:08	U	SMPL	
J4771-02LX5 - 1		-18	0.0205	-7470A	PPB	14 Sep 2018 15:40:16	U	SMPL	
J4771-02A - 1		14859	4.2424	-7470A	PPB	14 Sep 2018 15:42:24	U	SMPL	
J4882-02LX5 - 1		-196	-0.0300	-7470A	PPB	14 Sep 2018 15:44:31	U	SMPL	
J4882-02A - 1		14228	4.0634	-7470A	PPB	14 Sep 2018 15:46:40	U	SMPL	
J4882-01LX5 - 1		-171	-0.0229	-7470A	PPB	14 Sep 2018 15:48:47	U	SMPL	
J4882-01A - 1		16198	4.6224	-7470A	PPB	14 Sep 2018 15:50:55	U	SMPL	
CCV65 - 1	CCV65	17501	4.9922	-7470A	PPB	14 Sep 2018 15:53:03	U	SMPL	
CCB65 - 1	CCB65	-254	-0.0465	-7470A	PPB	14 Sep 2018 15:55:13	U	SMPL	