

21397923

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

9/14/2018 15:17

Test: Total CN ug/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.507	0.0	0.092	
ICB1	1.671	0.0	0.005	
CCV1	238.411	0.0	0.224	
CCB1	1.490	0.0	0.004	
PB112854BL	1.372	0.0	0.004	
PB112854BS	96.704	0.0	0.093	
LOWPB112854	10.828	0.0	0.013	
HIGHPB112854	491.325	0.0	0.459	
J4880-25	280.496	0.0	0.263	
CCV2	240.512	0.0	0.226	
CCB2	1.439	0.0	0.004	

N	11
Mean	132.705
SD	161.1357
CV%	121.42

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

9/14/2018 12:46


Test Total CN

Accepted 9/14/2018 12:46

Factor 1077
Bias 0.003

Coeff. of det. 0.999975

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	1.0859	0.0000	
2	5.0PPBCN	0.008	5.6775	5.0000	
3	10PPBCN	0.012	10.0267	10.0000	
4	50PPBCN	0.049	49.7746	50.0000	
5	100PPBCN	0.095	98.6379	100.0000	
6	250PPBCN	0.234	249.0193	250.0000	
7	500PPBCN	0.468	500.7780	500.0000	

Aquakem v. 7.2AQ1

Results from time period:

Fri Sep 14 15:14:44 2018

Fri Sep 14 15:17:32 2018

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	1.0859	µg/l	9/14/2018 12:30:36	
5.0PPBCN	A	Total CN	P	5.6775	µg/l	9/14/2018 12:30:37	
10PPBCN	A	Total CN	P	10.0267	µg/l	9/14/2018 12:30:38	
50PPBCN	A	Total CN	P	49.7746	µg/l	9/14/2018 12:30:39	
100PPBCN	A	Total CN	P	98.6379	µg/l	9/14/2018 12:30:40	
250PPBCN	A	Total CN	P	249.0193	µg/l	9/14/2018 12:30:41	
500PPBCN	A	Total CN	P	500.778	µg/l	9/14/2018 12:30:42	
ICV1	S	Total CN	P	95.5065	µg/l	9/14/2018 15:14:44	
ICB1	S	Total CN	P	1.6712	µg/l	9/14/2018 15:14:45	
CCV1	S	Total CN	P	238.4114	µg/l	9/14/2018 15:14:46	
CCB1	S	Total CN	P	1.4897	µg/l	9/14/2018 15:14:47	
PB112854BL	S	Total CN	P	1.372	µg/l	9/14/2018 15:14:48	
PB112854BS	S	Total CN	P	96.7039	µg/l	9/14/2018 15:14:49	
LOWPB112854	S	Total CN	P	10.8278	µg/l	9/14/2018 15:14:50	
HIGHPB112854	S	Total CN	P	491.325	µg/l	9/14/2018 15:14:51	
J4880-25	S	Total CN	P	280.4962	µg/l	9/14/2018 15:14:52	
CCV2	S	Total CN	P	240.5119	µg/l	9/14/2018 15:14:54	
CCB2	S	Total CN	P	1.4387	µg/l	9/14/2018 15:17:32	