


ISM02.4

LB 97926
INSTRUMENT ID: CV2

Linear

 μ Abs.:

42017



A= 0.0000e+000

B= 2.3892e-004

C= -6.2463e-003

Rho= 0.9997983

Accept=Accepted

*Slope
Y=1071000X*

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	O/D
0.00	0.000	0.019	0.019	105	0.000	105					—
0.2	0.200	0.192	-0.008	830	0.0 %	830					-4
2.5	2.500	2.443	-0.057	10253	0.0 %	10253					-2
5.0	5.000	5.125	0.125	21479	0.0 %	21479					3
7.5	7.500	7.388	-0.112	30949	0.0 %	30949					-1
10.0	10.000	10.032	0.032	42017	0.0 %	42017					0

*NB
9/11/18*

LB97926
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	105	-		0 ISM02.4	PPB	9/14/2018 16:13	Std
	0.2 S0.2	830	-		0.2 ISM02.4	PPB	9/14/2018 16:15	Std
	2.5 S2.5	10253	-		2.5 ISM02.4	PPB	9/14/2018 16:17	Std
	5 S5	21479	-		5 ISM02.4	PPB	9/14/2018 16:19	Std
	7.5 S7.5	30949	-		7.5 ISM02.4	PPB	9/14/2018 16:22	Std
	10 S10.0	42017	-		10 ISM02.4	PPB	9/14/2018 16:27	Std
ICV	ICV	16231	3.8716 -		ISM02.4	PPB	9/14/2018 16:30	SMPL
ICB	ICB	-257	-0.0676 -		ISM02.4	PPB	9/14/2018 16:32	SMPL
CCV066	CCV066	21598	5.1539 -		ISM02.4	PPB	9/14/2018 16:34	SMPL
CCB066	CCB066	-228	-0.0607 -		ISM02.4	PPB	9/14/2018 16:36	SMPL
PB112868BL	PBW003	-16	-0.0101 -		ISM02.4	PPB	9/14/2018 16:39	SMPL
J4785-01	MBEAQ2	29	0.0007 -		ISM02.4	PPB	9/14/2018 16:41	SMPL
J4785-02	MBEAQ2D	-2	-0.0067 -		ISM02.4	PPB	9/14/2018 16:43	SMPL
J4785-03	MBEAQ2S	4402	1.0455 -		ISM02.4	PPB	9/14/2018 16:45	SMPL
J4785-04	MBEAQ3	-46	-0.0172 -		ISM02.4	PPB	9/14/2018 16:48	SMPL
PB112869BL	PBW003	36	0.0024 -		ISM02.4	PPB	9/14/2018 16:50	SMPL
J4861-04	MA41S8	42607	10.1733 -		ISM02.4	PPB	9/14/2018 16:52	SMPL
J4861-05	MA41S8D	40139	9.5836 -		ISM02.4	PPB	9/14/2018 16:55	SMPL
J4861-06	MA41S8S	46426	11.0857 -		ISM02.4	PPB	9/14/2018 17:00	SMPL
PB112870BL	PBW003	-66	-0.022 -		ISM02.4	PPB	9/14/2018 17:05	SMPL
J4915-01	MCOAE9	58	0.0076 -		ISM02.4	PPB	9/14/2018 17:07	SMPL
J4915-02	MCOAF0	-21	-0.0113 -		ISM02.4	PPB	9/14/2018 17:09	SMPL
J4915-03	MCOAF1	61	0.0083 -		ISM02.4	PPB	9/14/2018 17:11	SMPL
J4915-04	MCOAF2	90	0.0153 -		ISM02.4	PPB	9/14/2018 17:14	SMPL
J4915-05	MCOAF4	64	0.009 -		ISM02.4	PPB	9/14/2018 17:16	SMPL
J4915-06	MCOAF5	96	0.0167 -		ISM02.4	PPB	9/14/2018 17:18	SMPL
J4915-07	MCOAF5D	8	-0.0043 -		ISM02.4	PPB	9/14/2018 17:21	SMPL
J4915-08	MCOAF5S	4330	1.0283 -		ISM02.4	PPB	9/14/2018 17:23	SMPL
CCV067	CCV067	20980	5.0062 -		ISM02.4	PPB	9/14/2018 17:25	SMPL
CCB067	CCB067	-122	-0.0354 -		ISM02.4	PPB	9/14/2018 17:27	SMPL
J4861-04DLX2	MA41S8	21060	5.0253 -		ISM02.4	PPB	9/14/2018 17:30	SMPL
J4861-05DLX2	MA41S8D	20314	4.8471 -		ISM02.4	PPB	9/14/2018 17:32	SMPL
J4861-06DLX2	MA41S8S	23603	5.6329 -		ISM02.4	PPB	9/14/2018 17:34	SMPL
CCV068	CCV068	21465	5.1221 -		ISM02.4	PPB	9/14/2018 17:37	SMPL
CCB068	CCB068	-114	-0.0335 -		ISM02.4	PPB	9/14/2018 17:39	SMPL