CHEMTECH

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB97932

Analyst By : sweta

Supervisor Review By : apatel

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.0	4.01	09/14/2018	18:00
2	CAL2	1	Water			21.9	7.01	09/14/2018	18:02
3	CAL3	1	Water			21.8	10.04	09/14/2018	18:04
4	ICV	1	Water			22.0	7.00	09/14/2018	18:06
5	CCV1	1	Water			21.8	2.00	09/14/2018	18:08
6	J4929-13	1	Water			22.3	6.34	09/14/2018	18:10
7	J4929-13DUP	1	Water			22.4	6.35	09/14/2018	18:12
8	CCV2	1	Water			21.9	12.00	09/14/2018	18:14

WORKLIST(Hardcopy Internal Chain)

LB97932

WorkList Name : pH091418 WorkList ID : 116749 Date : 9/14/2018 5:16:59 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/14/2018	Water	J4929-13	pH	Cool 4 deg C	PSEG01	I31	SP-PIT-11-BASELINE-WATEF	09/14/2018	9040C

Date/Time 09/14/18 5:50
Received by: SP
Relinquished by: CP

Date/Time 09/14/18 7:00
Received by: CP
Relinquished by: SP